

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102822\
 Data File : PS020917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2022 15:38
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 10:04:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 31 10:04:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.409	6.849	485.8E6	302.2E6	219.479	211.943
Target Compounds							
1) T	Dalapon	2.504	2.577	519.1E6	346.8E6	192.328	189.457
2) T	3,5-DICHL...	5.777	6.020	658.2E6	417.1E6	202.748	197.088
3) T	4-Nitroph...	6.244	6.456	306.6E6	170.1E6	196.080	192.742
5) T	DICAMBA	6.554	7.016	1909.0E6	1264.3E6	203.998	197.210
6) T	MCPD	6.708	7.111	123.0E6	88853969	18.930	18.924
7) T	MCPA	6.832	7.322	184.4E6	135.6E6	19.362	19.558
8) T	DICHLORPROP	7.154	7.643	496.3E6	314.8E6	205.565	198.700
9) T	2,4-D	7.353	7.933	603.2E6	384.4E6	204.719	199.300
10) T	Pentachlo...	7.588	8.366	7592.5E6	4881.8E6	215.170	211.102
11) T	2,4,5-TP ...	8.116	8.728	2969.0E6	1941.0E6	208.645	204.749
12) T	2,4,5-T	8.373	9.103	2968.4E6	1855.2E6	208.466	204.414
13) T	2,4-DB	8.882	9.611	582.3E6	209.4E6	203.410	191.533
14) T	DINOSEB	9.943	9.942	2380.0E6	1330.1E6	206.244	200.652
15) T	Picloram	9.768	10.910	4606.9E6	2637.5E6	206.080	200.553
16) T	DCPA	10.210	10.873	4209.6E6	2394.0E6	214.187	209.530

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102822\
Data File : PS020917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2022 15:38
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 10:04:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
Quant Title : 8080.M
QLast Update : Mon Oct 31 10:04:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

