

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102921\
 Data File : PS017277.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2021 12:50
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 23:41:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.155	6.736	703.0E6	318.7E6	459.917	545.231
Target Compounds							
2) T	3,5-DICHL...	6.419f	0.000	3321323	0	1.539	N.D. #
7) T	MCPA	7.641	0.000	6405085	0	1.038	N.D. #
15) T	Picloram	10.918	0.000	5008527	0	<MDL	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102921\
Data File : PS017277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2021 12:50
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 23:41:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.419 min
Delta R.T.: 0.043 min
Response: 3321323
Conc: 1.54 ng/ml

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 5.934 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.155 min
Delta R.T.: 0.004 min
Response: 702965392
Conc: 459.92 ng/ml



#4 2,4-DCAA


R.T.: 6.736 min
Delta R.T.: -0.001 min
Response: 318655613
Conc: 545.23 ng/ml



#7 MCPA

R.T.: 7.641 min
Delta R.T.: -0.018 min
Response: 6405085
Conc: 1.04 ug/ml

Instrument :
ECD_S
ClientSampleId :



#7 MCPA

R.T.: 0.000 min
Exp R.T. : 7.206 min
Response: 0
Conc: N.D.