

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102921\
 Data File : PS017278.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2021 13:14
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 23:41:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.151	6.737	1115.1E6	434.6E6	729.552	743.647
Target Compounds							
1) T	Dalapon	2.988	2.494	1463.2E6	444.8E6	673.830	687.452
2) T	3,5-DICHL...	6.375	5.935	1477.3E6	607.8E6	684.361	699.801
3) T	4-Nitroph...	6.981	6.362	556.7E6	207.4E6	638.283	643.276
5) T	DICAMBA	7.340	6.898	4193.3E6	1804.1E6	700.078	708.581
6) T	MCPD	7.509	6.994	318.5E6	145.4E6	71.240	72.182
7) T	MCPA	7.657	7.206	441.9E6	197.6E6	71.628	66.754
8) T	DICHLORPROP	8.031	7.514	1044.3E6	457.7E6	696.510	711.772
9) T	2,4-D	8.260	7.805	1224.8E6	534.9E6	691.490	710.727
10) T	Pentachlo...	8.564	8.229	14741.9E6	5880.9E6	710.520	720.287
11) T	2,4,5-TP ...	9.115	8.589	5796.9E6	2505.1E6	712.794	726.730
12) T	2,4,5-T	9.398	8.966	5859.5E6	2347.0E6	729.503	740.496
13) T	2,4-DB	9.948	9.470	1120.8E6	294.8E6	745.010	710.659
14) T	DINOSEB	11.100	9.795	4363.4E6	1755.1E6	770.269	795.408
15) T	Picloram	10.922	10.774	7646.0E6	2764.0E6	715.821	721.710
16) T	DCPA	11.368	10.725	7487.5E6	3111.5E6	735.210	729.147

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102921\
Data File : PS017278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2021 13:14
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 23:41:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

