

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102921\
 Data File : PS017282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Oct 2021 18:02
 Operator : DD\AJ
 Sample : M4068-07
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 23:42:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.153	6.739	378.5E6	168.3E6	247.607	287.913
Target Compounds							
1) T	Dalapon	2.988	2.495	193.3E6	62048369	89.013	95.895
2) T	3,5-DICHL...	6.380	5.939	206.4E6	77080647	95.619	88.748
3) T	4-Nitroph...	6.999	6.371	67572975	25542949	77.478	79.241
5) T	DICAMBA	7.341	6.899	553.6E6	229.4E6	92.419	90.099
6) T	MCPD	7.504	6.988	36218903	17426858	8.102	8.653
7) T	MCPA	7.650	7.197	60108374	31089623	9.743	10.502
8) T	DICHLORPROP	8.034	7.517	150.9E6	63972793	100.667	99.477
9) T	2,4-D	8.273	7.812	175.5E6	71967256	99.067	95.627
10) T	Pentachlo...	8.566	8.229	2121.5E6	850.5E6	102.252	104.173
11) T	2,4,5-TP ...	9.118	8.591	781.2E6	325.0E6	96.055	94.275
12) T	2,4,5-T	9.407	8.970	784.4E6	291.1E6	97.652	91.853
13) T	2,4-DB	9.959	9.476	134.7E6	37066323	89.544	89.365
14) T	DINOSEB	11.101	9.797	625.2E6	217.6E6	110.371	98.636
15) T	Picloram	10.933	10.781	1347.3E6	462.5E6	126.139	120.767
16) T	DCPA	11.368	10.726	1071.5E6	419.9E6	105.209	98.409

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102921\
Data File : PS017282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Oct 2021 18:02
Operator : DD\AJ
Sample : M4068-07
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
LOD-MDL-WATER-01-QT4-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 23:42:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

