

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103018\
 Data File : PS002208.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2018 22:27
 Operator : EI\SJ
 Sample : PB114301BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 PB114301BS

Manual Integrations
 APPROVED

Sohil
 10/31/2018 3:18:42 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 02:34:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102618.M
 Quant Title : 8080.M
 QLast Update : Sat Oct 27 01:34:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.443	10.933	178.7E6	183.5E6	567.885	510.915
Target Compounds							
1) T	Dalapon	2.958	2.396	211.0E6	278.1E6	460.057	616.007 #
2) T	3,5-DICHL...	9.873	9.088	226.9E6	209.7E6	499.873m	472.713
3) T	4-Nitroph...	11.034	10.077	97474725	74628735	584.279	418.532m#
5) T	DICAMBA	11.753	11.216	641.9E6	579.4E6	483.321	463.635
6) T	MCPD	12.093	11.445	44693194	42642671	45.242m	49.565
7) T	MCPA	12.341	11.819	57260568	81992050	38.135	60.582 #
8) T	DICHLORPROP	12.938	12.339	166.9E6	162.5E6	479.130	506.166
9) T	2,4-D	13.294	12.809	183.1E6	221.4E6	454.961	519.960
10) T	Pentachlo...	13.676	13.356	2453.2E6	2542.7E6	509.947	541.303
11) T	2,4,5-TP ...	14.542	13.974	949.9E6	1071.6E6	496.610m	585.688
12) T	2,4,5-T	14.938	14.537	875.3E6	776.6E6	488.388	518.605m
13) T	2,4-DB	15.684	15.188	108.8E6	116.5E6	463.412	525.883
14) T	DINOSEB	17.165	15.616	629.7E6	636.7E6	459.622	502.963
15) T	Picloram	16.929	16.793	1107.3E6	909.9E6	464.211	402.211
16) T	DCPA	17.508	16.846	708.1E6	1233.9E6	301.633m	521.105 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103018\
Data File : PS002208.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2018 22:27
Operator : EI\SJ
Sample : PB114301BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB114301BS

Manual Integrations
APPROVED

Sohil
10/31/2018 3:18:42 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 02:34:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102618.M
Quant Title : 8080.M
QLast Update : Sat Oct 27 01:34:10 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

