

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103019\
 Data File : PS007522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2019 19:59
 Operator : SM\AJ
 Sample : K5540-19MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleID :
 160-04-(0-1)MSD

Manual Integrations
 APPROVED

Ankita
 10/31/2019 10:20:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 00:47:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.449	5.176	326.7E6	157.8E6	238.732	224.303
Target Compounds							
1) T	Dalapon	2.029	1.742	386.9E6	204.1E6	231.246	224.237m
2) T	3,5-DICHL...	4.826	4.486	449.5E6	209.0E6	267.499	242.940
5) T	DICAMBA	5.599	5.300	1193.4E6	613.0E6	242.883	235.598
6) T	MCP	5.748	5.382	87620293	42080703	24.860	22.185
7) T	MCPA	5.878	5.554	130.0E6	84759549	25.214m	29.156
8) T	DICHLORPROP	6.221	5.826	322.7E6	169.9E6	275.325m	260.957m
9) T	2,4-D	6.435	6.095	412.0E6	229.4E6	334.311	287.545
10) T	Pentachlo...	6.693	6.458	665.1E6	338.5E6	37.253	35.197
11) T	2,4,5-TP ...	7.225	6.820	1773.1E6	942.2E6	254.397	247.718
12) T	2,4,5-T	7.496	7.178	1990.7E6	966.2E6	260.406	275.778
13) T	2,4-DB	8.023	7.664	293.6E6	120.4E6	307.908m	256.966
14) T	DINOSEB	9.107	7.972	258.2E6	131.1E6	44.372	46.701m
15) T	Picloram	8.940	8.907	2866.5E6	1474.2E6	259.709	240.092
16) T	DCPA	9.379	8.864	2417.1E6	915.3E6	269.935	182.524 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103019\
Data File : PS007522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2019 19:59
Operator : SM\AJ
Sample : K5540-19MSD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
160-04-(0-1)MSD

Manual Integrations
APPROVED

Ankita
10/31/2019 10:20:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 00:47:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

