

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103024\
 Data File : PS028079.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2024 13:14
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/31/2024
 Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 14:54:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS103024.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 30 14:45:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.055	7.595	665.1E6	283.2E6	221.373	215.585
Target Compounds							
1) T	Dalapon	2.527	2.588	711.2E6	443.8E6	202.013	190.157
2) T	3,5-DICHL...	6.255	6.573	904.6E6	394.4E6	204.235	193.259
3) T	4-Nitroph...	6.851	7.132	418.9E6	167.2E6	191.564	196.728
5) T	DICAMBA	7.232	7.787	2560.4E6	1135.6E6	198.546	192.032
6) T	MCP	7.406	7.888	160.3E6	91794789	17.355	19.878
7) T	MCPA	7.549	8.125	242.8E6	138.4E6	18.713	21.306m
8) T	DICHLORPROP	7.912	8.492	659.0E6	291.2E6	206.325	210.817
9) T	2,4-D	8.134	8.816	766.9E6	299.0E6	205.018	208.333
10) T	Pentachlo...	8.418	9.325	9884.6E6	3796.3E6	208.694	205.907
11) T	2,4,5-TP ...	8.982	9.705	3821.1E6	1470.9E6	204.719	207.518
12) T	2,4,5-T	9.266	10.120	4016.7E6	1383.5E6	204.445	206.699
13) T	2,4-DB	9.827	10.682	590.1E6	173.0E6	199.566	209.502
14) T	DINOSEB	10.997	11.056	2821.2E6	1002.8E6	200.143	207.754
15) T	Picloram	10.815	12.134	6191.7E6	1831.1E6	199.520	195.376
16) T	DCPA	11.296	12.089	4322.9E6	1715.8E6	204.075	212.221

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103024\
Data File : PS028079.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2024 13:14
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/31/2024
Supervised By : Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 14:54:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS103024.M
Quant Title : 8080.M
QLast Update : Wed Oct 30 14:45:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

