

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103024\
 Data File : PS028081.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2024 14:02
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 14:46:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS103024.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 30 14:45:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.055	7.596	2180.6E6	937.3E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.525	2.586	2325.9E6	1572.5E6	682.500	682.500
2) T	3,5-DICHL...	6.255	6.574	3003.5E6	1414.5E6	697.500	697.500
3) T	4-Nitroph...	6.850	7.131	1472.1E6	563.3E6	682.500	682.500
5) T	DICAMBA	7.232	7.789	8998.0E6	3994.9E6	705.000	705.000
6) T	MCPD	7.411	7.893	678.2E6	303.9E6	70.500	70.500
7) T	MCPA	7.554	8.132	913.4E6	396.1E6	69.750	69.750
8) T	DICHLORPROP	7.914	8.493	2190.3E6	863.1E6	705.000	705.000
9) T	2,4-D	8.135	8.817	2587.5E6	892.9E6	705.000	705.000
10) T	Pentachlo...	8.420	9.326	33338.2E6	11799.9E6	712.500	712.500
11) T	2,4,5-TP ...	8.984	9.705	13037.4E6	4472.1E6	712.500	712.500
12) T	2,4,5-T	9.268	10.120	13753.9E6	4266.5E6	712.500	712.500
13) T	2,4-DB	9.828	10.681	2093.6E6	518.0E6	712.500	712.500
14) T	DINOSEB	10.999	11.056	9772.8E6	3031.1E6	705.000	705.000
15) T	Picloram	10.816	12.132	22018.1E6	6089.6E6	712.500	712.500
16) T	DCPA	11.299	12.089	15084.5E6	5085.5E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103024\
Data File : PS028081.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2024 14:02
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 14:46:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS103024.M
Quant Title : 8080.M
QLast Update : Wed Oct 30 14:45:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

