

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103119\
 Data File : PS007527.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2019 09:16
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 11/1/2019 10:48:48 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 14:38:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.432	5.158	984.7E6	520.3E6	719.466	739.815
Target Compounds							
1) T	Dalapon	2.014	1.723	1117.7E6	580.7E6	667.997	638.123
2) T	3,5-DICHL...	4.808	4.467	1181.3E6	622.6E6	702.987	723.617
3) T	4-Nitroph...	5.285	4.848	403.2E6	250.6E6	671.393	634.584
5) T	DICAMBA	5.582	5.283	3646.8E6	1856.9E6	742.222	713.619
6) T	MCPD	5.734	5.367	281.2E6	129.3E6	79.774m	68.176m
7) T	MCPA	5.865	5.542	401.5E6	203.2E6	77.841	69.911
8) T	DICHLORPROP	6.204	5.809	897.4E6	466.3E6	765.511	716.010
9) T	2,4-D	6.416	6.076	993.7E6	559.6E6	806.338	701.371
10) T	Pentachlo...	6.676	6.440	13656.4E6	6934.9E6	764.971	721.152
11) T	2,4,5-TP ...	7.208	6.801	5377.8E6	2765.7E6	771.598	727.121
12) T	2,4,5-T	7.478	7.159	5740.1E6	2742.7E6	750.855	782.810m
13) T	2,4-DB	8.004	7.644	716.7E6	362.1E6	751.642	772.655
14) T	DINOSEB	9.089	7.953	4090.3E6	2028.9E6	702.900	722.693
15) T	Picloram	8.921	8.886	7801.1E6	4084.2E6	706.797	665.168
16) T	DCPA	9.362	8.846	6808.1E6	3702.7E6	760.318	738.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103119\
Data File : PS007527.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Oct 2019 09:16
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
11/1/2019 10:48:48 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 31 14:38:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

