

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103124\
 Data File : PS028149.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2024 06:54
 Operator : AR\AJ
 Sample : P4598-09MS
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-8MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/06/2024
 Supervised By : Ankita Jodhani 11/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 07:50:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS103124.M
 Quant Title : 8080.M
 QLast Update : Thu Oct 31 13:10:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.064	7.601	766.7E6	182.2E6	245.043m	182.368 #
Target Compounds							
1) T	Dalapon	2.526	2.589	856.4E6	417.5E6	208.553	193.670m
2) T	3,5-DICHL...	6.263	6.579	860.6E6	232.9E6	185.789	151.008
5) T	DICAMBA	7.242	7.794	2859.3E6	840.8E6	208.893m	184.834
6) T	MCPD	7.416	7.894	84043633	40890471	8.929m	13.062 #
7) T	MCPA	7.560	8.133	176.6E6	80137206	13.649	17.689m#
8) T	DICHLORPROP	7.925	8.500	614.6E6	175.9E6	180.705	160.861
9) T	2,4-D	8.148	8.824	1317.0E6	273.7E6	324.273	224.793m#
10) T	Pentachlo...	8.431	9.335	4031.0E6	1091.9E6	81.988	65.810
11) T	2,4,5-TP ...	8.996	9.715	2228.8E6	631.8E6	109.393	99.417
12) T	2,4,5-T	9.280	10.130	3156.5E6	807.4E6	145.611	131.907
13) T	2,4-DB	9.842	10.693	252.4E6	64546204	74.250m	85.331
14) T	DINOSEB	11.014	11.070	280.7E6	76972778	17.022	16.994
15) T	Picloram	10.831	12.142	7775.7E6	1755.2E6	225.331	212.087
16) T	DCPA	11.314	12.099	7807.4E6	2447.5E6	326.988	347.738

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS103124\
Data File : PS028149.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2024 06:54
Operator : AR\AJ
Sample : P4598-09MS
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-8MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/06/2024
Supervised By : Ankita Jodhani 11/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 07:50:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS103124.M
Quant Title : 8080.M
QLast Update : Thu Oct 31 13:10:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

