

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110119\
 Data File : PS007566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Nov 2019 14:29
 Operator : SM\AJ
 Sample : PB124480BSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB124480BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 01 15:29:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 14 13:54:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.451	5.177	730.6E6	355.9E6	533.775	506.056
Target Compounds							
1) T	Dalapon	2.033	1.743	794.8E6	511.2E6	475.003	561.716
2) T	3,5-DICHL...	4.827	4.486	860.4E6	421.8E6	512.004	490.173
3) T	4-Nitroph...	5.306	4.868	289.7E6	167.9E6	482.390	425.141
5) T	DICAMBA	5.600	5.301	2527.9E6	1275.9E6	514.486	490.342
6) T	MCPD	5.751	5.384	181.1E6	85673021	51.377	45.167
7) T	MCPA	5.882	5.558	269.9E6	142.8E6	52.334	49.127
8) T	DICHLORPROP	6.222	5.828	635.5E6	325.8E6	542.129	500.283
9) T	2,4-D	6.435	6.096	706.9E6	390.9E6	573.614	490.032
10) T	Pentachlo...	6.694	6.459	10011.7E6	5033.5E6	560.813	523.429
11) T	2,4,5-TP ...	7.226	6.820	3893.4E6	1966.6E6	558.616	517.045
12) T	2,4,5-T	7.496	7.179	4207.0E6	1901.9E6	550.316	542.850
13) T	2,4-DB	8.023	7.664	558.8E6	254.3E6	586.090	542.634
14) T	DINOSEB	9.108	7.972	2976.4E6	1427.5E6	511.480	508.467
15) T	Picloram	8.939	8.906	5370.6E6	2717.9E6	486.583	442.644
16) T	DCPA	9.380	8.864	5096.9E6	2669.8E6	569.216	532.392

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110119\
Data File : PS007566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2019 14:29
Operator : SM\AJ
Sample : PB124480BSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB124480BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 01 15:29:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101419.M
Quant Title : 8080.M
QLast Update : Mon Oct 14 13:54:23 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

