

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS110122\  
 Data File : PS020974.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 01 Nov 2022 09:58  
 Operator : AR\AJ  
 Sample : HSTDCCC750  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 HSTDCCC750

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 01 13:49:17 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS102822.M  
 Quant Title : 8080.M  
 QLast Update : Mon Oct 31 10:07:57 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1    | Resp#2    | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|-----------|-----------|---------|---------|
| -----                       |              |        |        |           |           |         |         |
| System Monitoring Compounds |              |        |        |           |           |         |         |
| 4) S                        | 2,4-DCAA     | 6.410  | 6.848  | 1569.4E6  | 1064.6E6  | 751.981 | 769.662 |
| Target Compounds            |              |        |        |           |           |         |         |
| 1) T                        | Dalapon      | 2.507  | 2.579  | 1811.3E6  | 1253.7E6  | 684.414 | 689.340 |
| 2) T                        | 3,5-DICHL... | 5.777  | 6.019  | 2131.6E6  | 1477.4E6  | 696.096 | 723.915 |
| 3) T                        | 4-Nitroph... | 6.246  | 6.456  | 1040.1E6  | 621.7E6   | 691.503 | 723.130 |
| 5) T                        | DICAMBA      | 6.555  | 7.015  | 6276.9E6  | 4523.6E6  | 710.483 | 728.430 |
| 6) T                        | MCPD         | 6.714  | 7.115  | 466.4E6   | 343.2E6   | 71.459  | 72.234  |
| 7) T                        | MCPA         | 6.841  | 7.328  | 656.9E6   | 486.7E6   | 70.010  | 71.723  |
| 8) T                        | DICHLORPROP  | 7.154  | 7.640  | 1603.9E6  | 1115.0E6  | 702.651 | 724.083 |
| 9) T                        | 2,4-D        | 7.354  | 7.931  | 1973.7E6  | 1368.0E6  | 706.200 | 730.432 |
| 10) T                       | Pentachlo... | 7.589  | 8.365  | 22422.8E6 | 15360.8E6 | 711.630 | 724.158 |
| 11) T                       | 2,4,5-TP ... | 8.116  | 8.725  | 9523.9E6  | 6556.8E6  | 716.908 | 730.652 |
| 12) T                       | 2,4,5-T      | 8.374  | 9.101  | 9634.0E6  | 6335.6E6  | 724.001 | 734.815 |
| 13) T                       | 2,4-DB       | 8.882  | 9.609  | 1933.8E6  | 894.8E6   | 700.901 | 803.898 |
| 14) T                       | DINOSEB      | 9.944  | 9.940  | 7626.8E6  | 4630.7E6  | 702.452 | 723.402 |
| 15) T                       | Picloram     | 9.772  | 10.911 | 15302.4E6 | 9221.0E6  | 724.449 | 726.777 |
| 16) T                       | DCPA         | 10.211 | 10.870 | 12804.0E6 | 7807.9E6  | 707.508 | 731.224 |
| -----                       |              |        |        |           |           |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS110122\  
Data File : PS020974.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 01 Nov 2022 09:58  
Operator : AR\AJ  
Sample : HSTDCCC750  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :  
HSTDCCC750

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 01 13:49:17 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS102822.M  
Quant Title : 8080.M  
QLast Update : Mon Oct 31 10:07:57 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

