

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110218\
 Data File : PS002281.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2018 23:23
 Operator : EI\SJ
 Sample : J5755-04MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-1-SMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 06:39:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102618.M
 Quant Title : 8080.M
 QLast Update : Sat Oct 27 01:34:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.451	10.942	109.7E6	107.6E6	348.565	299.622
Target Compounds							
1) T	Dalapon	2.973	2.411	191.5E6	227.4E6	417.588	503.558
2) T	3,5-DICHL...	9.881	9.098	201.0E6	181.5E6	442.903	409.221
3) T	4-Nitroph...	11.100	10.099	4454715	11443234	26.702	64.176 #
5) T	DICAMBA	11.760	11.225	621.0E6	593.7E6	467.627	475.091
6) T	MCPD	12.100	11.454	59203956	47622665	59.931m	55.353
7) T	MCPA	12.349	11.829	71330209	74074944	47.506	54.732
8) T	DICHLORPROP	12.946	12.348	175.8E6	166.7E6	504.948	519.114m
9) T	2,4-D	13.303	12.818	216.7E6	236.8E6	538.527	556.099
10) T	Pentachlo...	13.684	13.365	1823.9E6	1852.1E6	379.130	394.275
11) T	2,4,5-TP ...	14.551	13.983	976.3E6	1064.5E6	510.369	581.844
12) T	2,4,5-T	14.947	14.546	941.0E6	628.7E6	525.040	419.865m
13) T	2,4-DB	15.693	15.196	129.1E6	125.6E6	549.923	567.091
14) T	DINOSEB	17.174	15.625	423.5E6	419.8E6	309.088	331.604
15) T	Picloram	16.939	16.802	1102.3E6	944.7E6	462.122	417.584
16) T	DCPA	17.516	16.802	852.3E6	944.7E6	363.054m	398.953

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110218\
Data File : PS002281.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2018 23:23
Operator : EI\SJ
Sample : J5755-04MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TP-1-SMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 06:39:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102618.M
Quant Title : 8080.M
QLast Update : Sat Oct 27 01:34:10 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

