

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110221\
 Data File : PS017305.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2021 17:08
 Operator : DD\AJ
 Sample : M4395-01MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-FFSMS

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/03/2021
 Supervised By :mohammad ahmed 11/04/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 02:35:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.147	6.735	768.2E6	319.1E6	502.594m	546.006m
Target Compounds							
1) T	Dalapon	2.989	2.495	581.9E6	156.7E6	267.976	242.190
2) T	3,5-DICHL...	6.374	5.934	660.1E6	246.0E6	305.808	283.207m
3) T	4-Nitroph...	6.979	6.362	235.5E6	83049009	270.070	257.640m
5) T	DICAMBA	7.337	6.895	1753.8E6	736.6E6	292.802m	289.299m
6) T	MCP	7.503	6.988	132.6E6	58193899	29.654	28.897
7) T	MCPA	7.651	7.198	195.5E6	98096867	31.687	33.137
8) T	DICHLORPROP	8.029	7.514	498.2E6	186.3E6	332.250	289.764
9) T	2,4-D	8.258	7.804	570.3E6	256.6E6	322.004	340.917
10) T	Pentachlo...	8.563	8.227	6193.0E6	2466.3E6	298.485	302.067
11) T	2,4,5-TP ...	9.112	8.587	2459.0E6	1003.0E6	302.359	290.960m
12) T	2,4,5-T	9.396	8.965	2343.3E6	971.0E6	291.736	306.348
13) T	2,4-DB	9.947	9.469	330.7E6	109.1E6	219.844	263.151
14) T	DINOSEB	11.096	9.794	275.0E6	134.7E6	48.543	61.035 #
15) T	Picloram	10.920	10.773	3607.9E6	1288.6E6	337.770	336.468
16) T	DCPA	11.365	10.723	3006.6E6	1501.5E6	295.223	351.869

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110221\
 Data File : PS017305.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2021 17:08
 Operator : DD\AJ
 Sample : M4395-01MS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-FFSMS

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/03/2021
 Supervised By :mohammad ahmed 11/04/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 02:35:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

