

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110221\
 Data File : PS017306.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2021 17:32
 Operator : DD\AJ
 Sample : M4395-01MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-FFSMSD

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/03/2021
 Supervised By :mohammad ahmed 11/04/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 02:36:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.148	6.735	724.1E6	302.5E6	473.723m	517.524m
Target Compounds							
1) T	Dalapon	2.988	2.495	584.0E6	158.0E6	268.963	244.114
2) T	3,5-DICHL...	6.373	5.935	665.9E6	251.3E6	308.463	289.313m
3) T	4-Nitroph...	6.979	6.362	237.9E6	84399638	272.790	261.830m
5) T	DICAMBA	7.336	6.896	1762.4E6	736.5E6	294.245m	289.289m
6) T	MCPD	7.503	6.989	132.4E6	58556158	29.621	29.076
7) T	MCPA	7.650	7.198	196.1E6	98725905	31.784	33.349
8) T	DICHLORPROP	8.028	7.513	501.9E6	187.9E6	334.764	292.146
9) T	2,4-D	8.258	7.804	568.3E6	259.7E6	320.829	345.059
10) T	Pentachlo...	8.562	8.227	6218.2E6	2480.0E6	299.702	303.742
11) T	2,4,5-TP ...	9.112	8.587	2474.7E6	1011.3E6	304.287	293.370m
12) T	2,4,5-T	9.395	8.965	2360.4E6	980.7E6	293.874	309.427
13) T	2,4-DB	9.947	9.469	334.5E6	111.1E6	222.342	267.852
14) T	DINOSEB	11.096	9.794	275.0E6	136.8E6	48.551	61.997 #
15) T	Picloram	10.920	10.773	3650.2E6	1266.9E6	341.729	330.814
16) T	DCPA	11.366	10.723	3029.0E6	1499.5E6	297.421	351.401

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110221\
Data File : PS017306.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2021 17:32
Operator : DD\AJ
Sample : M4395-01MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-FFSMSD

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/03/2021
Supervised By :mohammad ahmed 11/04/2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 02:36:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

