

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110221\
 Data File : PS017321.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2021 23:34
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/03/2021
 Supervised By :mohammad ahmed 11/04/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 02:43:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.150	6.738	1166.6E6	446.2E6	763.254	763.440
Target Compounds							
1) T	Dalapon	2.988	2.492	1598.0E6	482.7E6	735.932	745.993
2) T	3,5-DICHL...	6.374	5.936	1498.5E6	626.1E6	694.171	720.901
3) T	4-Nitroph...	6.980	6.363	557.7E6	219.4E6	639.464	680.587
5) T	DICAMBA	7.338	6.898	4278.7E6	1869.2E6	714.341m	734.138
6) T	MCPD	7.508	6.995	324.8E6	149.9E6	72.665	74.447
7) T	MCPA	7.656	7.206	453.0E6	209.0E6	73.421	70.614
8) T	DICHLORPROP	8.029	7.515	1064.9E6	470.2E6	710.207	731.174
9) T	2,4-D	8.258	7.806	1277.2E6	550.9E6	721.095	731.961
10) T	Pentachlo...	8.563	8.229	14997.4E6	6060.1E6	722.835	742.230
11) T	2,4,5-TP ...	9.112	8.590	5892.5E6	2576.5E6	724.542	747.427
12) T	2,4,5-T	9.397	8.966	5985.8E6	2402.2E6	745.237	757.921
13) T	2,4-DB	9.946	9.471	1162.7E6	301.3E6	772.895	726.513
14) T	DINOSEB	11.097	9.796	4414.3E6	1778.1E6	779.249	805.814
15) T	Picloram	10.920	10.774	8409.9E6	3068.1E6	787.339	801.133
16) T	DCPA	11.365	10.725	7606.5E6	3202.5E6	746.892	750.485

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110221\
 Data File : PS017321.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2021 23:34
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations APPROVED

Reviewed By :Ankita Jodhani 11/03/2021
 Supervised By :mohammad ahmed 11/04/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 02:43:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

