

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110321\
 Data File : PS017324.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Nov 2021 09:56
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 03 14:57:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.150	6.737	1134.5E6	447.1E6	742.249	765.002
Target Compounds							
1) T	Dalapon	2.988	2.493	1582.8E6	479.3E6	728.918	740.734
2) T	3,5-DICHL...	6.374	5.935	1505.8E6	628.2E6	697.527	723.276
3) T	4-Nitroph...	6.980	6.363	552.6E6	217.5E6	633.640	674.704
5) T	DICAMBA	7.339	6.898	4302.5E6	1875.6E6	718.315	736.683
6) T	MCPD	7.508	6.995	325.8E6	151.0E6	72.873	75.000
7) T	MCPA	7.656	7.206	453.8E6	208.3E6	73.563	70.362
8) T	DICHLORPROP	8.029	7.515	1068.2E6	471.1E6	712.440	732.523
9) T	2,4-D	8.259	7.805	1266.0E6	547.2E6	714.757	727.131
10) T	Pentachlo...	8.563	8.228	15063.6E6	6062.4E6	726.025	742.521
11) T	2,4,5-TP ...	9.113	8.589	5917.6E6	2582.0E6	727.636	749.013
12) T	2,4,5-T	9.397	8.966	5960.5E6	2394.9E6	742.076	755.607
13) T	2,4-DB	9.946	9.470	1134.6E6	297.1E6	754.230	716.183
14) T	DINOSEB	11.097	9.796	4418.9E6	1773.3E6	780.057	803.634
15) T	Picloram	10.921	10.774	8324.1E6	3022.8E6	779.307	789.295
16) T	DCPA	11.365	10.724	7638.9E6	3200.1E6	750.076	749.910

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110321\
Data File : PS017324.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Nov 2021 09:56
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 03 14:57:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

