

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS110323.M
 Title : 8080.M
 Last Update : Fri Nov 03 15:07:01 2023
 Response Via : Initial Calibration

Calibration Files

750 =PS024755.D 1000 =PS024756.D 1500 =PS024757.D
 500 =PS024754.D 200 =PS024753.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	2.384	2.343	2.337	2.439	2.541	2.409	E6	3.51
2) T	3,5-DICHLOROBENZOI...	3.105	3.007	2.900	3.223	3.450	3.137	E6	6.75
3) T	4-Nitrophenol	1.282	1.271	1.278	1.307	1.343	1.296	E6	2.26
4) S	2,4-DCAA	2.007	1.946	1.888	2.091	2.276	2.042	E6	7.39
5) T	DICAMBA	8.949	8.709	8.414	9.197	9.469	8.948	E6	4.59
6) T	MCPP	6.348	6.535	6.677	6.143	5.447	6.230	E6	7.73
7) T	MCPA	8.903	8.930	9.041	8.775	8.217	8.773	E6	3.70
8) T	DICHLORPROP	2.300	2.235	2.175	2.392	2.596	2.340	E6	7.03
9) T	2,4-D	2.708	2.632	2.564	2.814	3.027	2.749	E6	6.60
10) T	Pentachlorophenol	3.326	3.169	2.589	3.506	3.699	3.258	E7	12.99
11) T	2,4,5-TP (SILVEX)	1.385	1.336	1.271	1.444	1.516	1.390	E7	6.83
12) T	2,4,5-T	1.303	1.261	1.210	1.355	1.409	1.308	E7	5.97
13) T	2,4-DB	2.093	2.065	2.044	2.135	2.178	2.103	E6	2.58
14) T	DINOSEB	0.976	0.948	0.927	1.016	1.064	0.986	E7	5.53
15) T	Picloram	1.784	1.764	1.749	1.802	1.795	1.779	E7	1.24
16) T	DCPA	1.694	1.622	1.530	1.786	1.925	1.711	E7	8.89

Signal #2 Calibration Files

750 =PS024755.D 1000 =PS024756.D 1500 =PS024757.D
 500 =PS024754.D 200 =PS024753.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	1.042	1.033	1.036	1.057	1.099	1.054	E6	2.56
2) T	3,5-DICHLOROBENZOI...	1.072	1.055	1.033	1.088	1.106	1.071	E6	2.64
3) T	4-Nitrophenol	4.336	4.278	4.261	4.443	4.636	4.391	E5	3.51
4) S	2,4-DCAA	6.712	6.575	6.432	6.849	7.162	6.746	E5	4.15
5) T	DICAMBA	2.936	2.896	2.839	2.951	2.906	2.906	E6	1.50
6) T	MCPP	1.987	2.013	2.052	1.932	1.717	1.940	E6	6.81
7) T	MCPA	3.068	3.028	3.022	3.094	3.204	3.083	E6	2.39
8) T	DICHLORPROP	7.267	7.126	6.984	7.415	7.733	7.305	E5	3.94
9) T	2,4-D	8.911	8.781	8.650	9.078	9.336	8.951	E5	2.98
10) T	Pentachlorophenol	1.074	1.038	0.982	1.108	1.111	1.063	E7	5.08
11) T	2,4,5-TP (SILVEX)	4.142	4.052	3.935	4.218	4.205	4.110	E6	2.88
12) T	2,4,5-T	4.024	3.958	3.866	4.081	4.022	3.990	E6	2.06
13) T	2,4-DB	4.895	4.910	4.960	4.868	4.745	4.876	E5	1.65
14) T	DINOSEB	2.854	2.801	2.790	2.908	2.902	2.851	E6	1.93
15) T	Picloram	6.156	6.172	6.175	6.096	5.780	6.076	E6	2.77
16) T	DCPA	5.030	4.854	4.648	5.238	5.359	5.026	E6	5.70

(#) = Out of Range