

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110325\
 Data File : PS032260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2025 00:39
 Operator : AR\AJ
 Sample : Q3516-01MSD
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HD-01-103125MSD

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/04/2025
 Supervised By :Yogesh Patel 11/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 04 05:43:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.284	7.624	193.5E6	244.6E6	42.708	67.101 #
Target Compounds							
1) T	Dalapon	2.642	2.605	428.6E6	3182.4E6	59.700m	457.639m#
2) T	3,5-DICHL...	6.454	6.595	177.5E6	122.7E6	27.452	20.420m#
5) T	DICAMBA	7.467	7.809	739.1E6	707.0E6	39.178	40.930m
6) T	MCP	7.682f	7.902	55780332	3695217	5.122m	<MDL #
7) T	MCPA	7.792	8.147	3738194	25969612	<MDL	1.593m#
8) T	DICHLORPROP	8.178	8.527	116.0E6	131.7E6	26.081m	27.622
9) T	2,4-D	8.393	8.882	369.1E6	215.0E6	81.813m	40.980m#
10) T	Pentachlo...	8.704	9.359	882.1E6	940.7E6	14.059	16.025
11) T	2,4,5-TP ...	9.284	9.744	459.2E6	708.3E6	17.653	26.377m#
12) T	2,4,5-T	9.587	10.187	458.2E6	602.5E6	19.220m	25.401m#
13) T	2,4-DB	10.182f	10.768f	47078233	15436719	17.866m	4.093m#
14) T	DINOSEB	11.345	11.114	88277077	120.4E6	4.243m	6.805 #
15) T	Picloram	11.204	12.242	1268.3E6	1920.5E6	59.752	51.651
16) T	DCPA	11.646	12.140	3493.8E6	3521.6E6	104.129	110.531

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110325\
Data File : PS032260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2025 00:39
Operator : AR\AJ
Sample : Q3516-01MSD
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HD-01-103125MSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/04/2025
Supervised By :Yogesh Patel 11/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 05:43:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

