

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110420\
 Data File : PS013111.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2020 17:40
 Operator : DD\AJ
 Sample : L4600-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 00:54:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.209	6.913	337.0E6	169.3E6	376.951	435.063
Target Compounds							
1) T	Dalapon	2.995	2.590	70739112	164.7E6	66.860	310.742 #
2) T	3,5-DICHL...	0.000	6.095	0	45113092	N.D.	73.334
3) T	4-Nitroph...	7.098f	6.575f	11601925	1773300	27.204	7.446 #
5) T	DICAMBA	7.385	7.046	7631913	6346374	2.137	3.823 #
7) T	MCPA	7.712	0.000	43849479	0	11.484	N.D. #
8) T	DICHLORPROP	0.000	7.746f	0	11325103	N.D.	25.338
10) T	Pentachlo...	8.682f	0.000	8910750	0	<MDL	N.D. #
11) T	2,4,5-TP ...	0.000	8.740f	0	21309118	N.D.	9.195
15) T	Picloram	10.974	0.000	15515447	0	2.360	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110420\
Data File : PS013111.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2020 17:40
Operator : DD\AJ
Sample : L4600-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 00:54:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.995 min
Delta R.T.: -0.037 min
Response: 70739112
Conc: 66.86 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.590 min
Delta R.T.: -0.033 min
Response: 164685819
Conc: 310.74 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.427 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.095 min
Delta R.T.: 0.028 min
Response: 45113092
Conc: 73.33 ng/ml



#3 4-Nitrophenol


R.T.: 7.098 min
Delta R.T.: 0.053 min
Response: 11601925
Conc: 27.20 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.575 min
Delta R.T.: 0.048 min
Response: 1773300
Conc: 7.45 ng/ml




#4 2,4-DCAA

R.T.: 7.209 min
Delta R.T.: -0.001 min
Response: 336967055
Conc: 376.95 ng/ml



#4 2,4-DCAA


R.T.: 6.913 min
Delta R.T.: 0.000 min
Response: 169278439
Conc: 435.06 ng/ml



#5 DICAMBA


R.T.: 7.385 min
Delta R.T.: -0.015 min
Response: 7631913
Conc: 2.14 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.046 min
Delta R.T.: -0.034 min
Response: 6346374
Conc: 3.82 ng/ml




#7 MCPA

R.T.: 7.712 min
Delta R.T.: -0.004 min
Response: 43849479
Conc: 11.48 ug/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 7.389 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.092 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.746 min
Delta R.T.: 0.040 min
Response: 11325103
Conc: 25.34 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 9.179 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 8.740 min
Delta R.T.: -0.049 min
Response: 21309118
Conc: 9.19 ng/ml



#15 Picloram

R.T.: 10.974 min
Delta R.T.: -0.032 min
Response: 15515447
Conc: 2.36 ng/ml

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 10.999 min
Response: 0
Conc: N.D.