

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110424\
 Data File : PS028184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2024 21:11
 Operator : AR\AJ
 Sample : P4578-06MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WB-305-BOTMSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/05/2024
 Supervised By : Ankita Jodhani 11/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 05 03:55:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110424.M
 Quant Title : 8080.M
 QLast Update : Mon Nov 04 23:43:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.606f	7.902	753.8E6	486.1E6	298.255m	290.183
Target Compounds							
1) T	Dalapon	2.976f	2.785	902.1E6	1010.7E6	282.171	359.477 #
2) T	3,5-DICHL...	6.752f	6.834	1385.0E6	1035.3E6	412.317m	407.654
3) T	4-Nitroph...	7.491	7.419	22751514	3316904	16.960m	2.841m#
5) T	DICAMBA	7.804f	8.108	3777.4E6	3081.0E6	398.063m	405.267
6) T	MCP	8.031	8.457	31015862	268.4E6	5.104	38.606 #
7) T	MCPA	8.187	8.457	11352630	268.4E6	1.387m	38.195 #
8) T	DICHLORPROP	8.527f	8.834	1032.8E6	809.3E6	439.752m	432.682
9) T	2,4-D	8.862f	9.171	126.3E6	1011.7E6	48.781m	484.763 #
10) T	Pentachlo...	9.081f	9.713	11307.6E6	8603.0E6	329.997m	308.932
11) T	2,4,5-TP ...	9.660f	10.093	6777.6E6	8620.6E6	523.814	768.909 #
12) T	2,4,5-T	9.956f	10.516	6533.4E6	4902.6E6	511.062	441.813
13) T	2,4-DB	10.538f	11.087	869.5E6	570.5E6	447.663m	424.957m
14) T	DINOSEB	11.769f	11.473	2969.4E6	1914.2E6	341.241	244.268 #
15) T	Picloram	11.572f	12.584	9070.2E6	6041.2E6	524.726m	377.440 #
16) T	DCPA	12.060f	12.528	9270.7E6	6334.7E6	645.339	514.069

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110424\
Data File : PS028184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2024 21:11
Operator : AR\AJ
Sample : P4578-06MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WB-305-BOTMSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/05/2024
Supervised By : Ankita Jodhani 11/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 05 03:55:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110424.M
Quant Title : 8080.M
QLast Update : Mon Nov 04 23:43:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

