

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110524\
 Data File : PS028223.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Nov 2024 19:48
 Operator : AR\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS110524

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 10:44:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110524.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 06 10:44:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.265	7.768	1293.1E6	1271.7E6	717.173	813.143
Target Compounds							
1) T	Dalapon	2.657	2.711	2160.9E6	1966.5E6	678.321	710.529
2) T	3,5-DICHL...	6.433	6.718	1882.0E6	1764.2E6	671.499	752.963
3) T	4-Nitroph...	7.064	7.294	771.6E6	773.1E6	651.024	730.305
5) T	DICAMBA	7.455	7.970	5461.6E6	5645.2E6	691.464	782.664
6) T	MCPD	7.639	8.072	389.1E6	420.7E6	70.650	81.304
7) T	MCPA	7.789	8.318	557.8E6	563.9E6	68.773	78.612
8) T	DICHLORPROP	8.166	8.687	1215.7E6	1345.7E6	672.907	773.091
9) T	2,4-D	8.397	9.018	1368.7E6	1499.5E6	675.077	769.013
10) T	Pentachlo...	8.698	9.550	18583.1E6	20213.7E6	694.799	781.351
11) T	2,4,5-TP ...	9.280	9.927	6943.0E6	8395.3E6	693.655	780.793
12) T	2,4,5-T	9.573	10.348	7091.5E6	8285.3E6	694.083	774.461
13) T	2,4-DB	10.151	10.916	1145.0E6	1046.1E6	694.312	754.695
14) T	DINOSEB	11.368	11.296	5127.3E6	5676.5E6	682.410	756.730
15) T	Picloram	11.172	12.397	12421.0E6	12429.1E6	711.830	757.057
16) T	DCPA	11.660	12.344	8789.9E6	9364.1E6	706.026	742.960

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110524\
Data File : PS028223.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Nov 2024 19:48
Operator : AR\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS110524

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 10:44:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110524.M
Quant Title : 8080.M
QLast Update : Wed Nov 06 10:44:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

