

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110624\
 Data File : PS028254.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2024 10:12
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/08/2024
 Supervised By : Ankita Jodhani 11/08/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 11:33:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110624.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 06 11:33:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.262	7.765	1265.5E6	870.6E6	492.828m	501.380
Target Compounds							
1) T	Dalapon	2.654	2.710	1576.6E6	1334.9E6	449.078	454.559
2) T	3,5-DICHL...	6.431	6.716	1771.1E6	1178.0E6	470.764	465.677
3) T	4-Nitroph...	7.062	7.291	748.3E6	518.6E6	456.609	456.042
5) T	DICAMBA	7.452	7.967	5332.7E6	3769.1E6	471.139	466.034
6) T	MCPD	7.634	8.067	349.6E6	280.7E6	45.882	46.335
7) T	MCPA	7.784	8.312	480.5E6	382.5E6	45.988	46.343
8) T	DICHLORPROP	8.163	8.684	1341.1E6	943.2E6	475.774	470.524
9) T	2,4-D	8.394	9.016	1503.2E6	1049.6E6	475.477	470.744
10) T	Pentachlo...	8.694	9.546	21029.5E6	14322.7E6	481.316	479.818
11) T	2,4,5-TP ...	9.277	9.923	8261.8E6	5990.7E6	479.437	475.627
12) T	2,4,5-T	9.570	10.344	8407.1E6	5907.4E6	479.082	475.572
13) T	2,4-DB	10.147	10.912	1342.8E6	743.8E6	475.218	471.354
14) T	DINOSEB	11.365	11.293	6850.2E6	4122.6E6	470.326	469.215
15) T	Picloram	11.169	12.393	13862.1E6	8612.2E6	472.396	469.122
16) T	DCPA	11.657	12.339	11638.0E6	6852.1E6	482.282	480.529

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110624\
Data File : PS028254.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2024 10:12
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/08/2024
Supervised By : Ankita Jodhani 11/08/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 11:33:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110624.M
Quant Title : 8080.M
QLast Update : Wed Nov 06 11:33:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

