

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110725\
 Data File : PS032310.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2025 02:37
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/10/2025
 Supervised By :Yogesh Patel 11/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 05:50:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.283	7.617	3252.8E6	2802.8E6	717.853m	768.845
Target Compounds							
1) T	Dalapon	2.656	2.605	4891.8E6	4708.7E6	681.430	677.115
2) T	3,5-DICHL...	6.450	6.586	4686.6E6	4348.4E6	724.780	723.663
3) T	4-Nitroph...	7.106	7.186	959.0E6	1709.5E6	714.382	753.113
5) T	DICAMBA	7.466	7.808	13834.6E6	12656.8E6	733.300	732.774
6) T	MCPD	7.642	7.905	811.4E6	1001.3E6	74.511	75.283
7) T	MCPA	7.791	8.150	979.0E6	1202.4E6	72.920	73.749
8) T	DICHLORPROP	8.173	8.522	3137.5E6	3446.8E6	705.261	722.721
9) T	2,4-D	8.413	8.865	3228.7E6	3840.5E6	715.620	732.134
10) T	Pentachlo...	8.703	9.358	47941.9E6	44172.8E6	764.146	752.484
11) T	2,4,5-TP ...	9.279	9.745	19089.8E6	19634.7E6	733.897	731.168
12) T	2,4,5-T	9.579	10.173	17713.8E6	17573.9E6	743.083	740.949
13) T	2,4-DB	10.156	10.738	1933.0E6	2754.7E6	733.540	730.425
14) T	DINOSEB	11.345	11.105	15062.1E6	12718.5E6	723.877	719.071
15) T	Picloram	11.186	12.226	15551.9E6	25899.0E6	732.696	696.544
16) T	DCPA	11.645	12.140	24772.4E6	23355.5E6	738.326	733.054

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110725\
Data File : PS032310.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2025 02:37
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/10/2025
Supervised By :Yogesh Patel 11/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 05:50:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

