

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110822\
 Data File : PS021082.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Nov 2022 17:36
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 00:40:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 31 10:07:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.409	6.846	1632.4E6	1111.6E6	782.149	803.651
Target Compounds							
1) T	Dalapon	2.508	2.578	1861.8E6	1324.6E6	703.471	728.273
2) T	3,5-DICHL...	5.776	6.017	2187.1E6	1573.5E6	714.199	770.958
3) T	4-Nitroph...	6.247	6.456	1002.2E6	616.9E6	666.356	717.517
5) T	DICAMBA	6.553	7.012	6489.7E6	4770.5E6	734.564	768.179
6) T	MCPD	6.712	7.112	484.9E6	364.4E6	74.292	76.711
7) T	MCPA	6.839	7.326	676.0E6	508.6E6	72.042	74.962
8) T	DICHLORPROP	7.153	7.638	1644.1E6	1173.4E6	720.256	762.021
9) T	2,4-D	7.354	7.929	1981.9E6	1400.8E6	709.139	747.952
10) T	Pentachlo...	7.586	8.360	23171.3E6	16253.1E6	735.384	766.225
11) T	2,4,5-TP ...	8.114	8.721	9615.6E6	6868.6E6	723.808	765.405
12) T	2,4,5-T	8.373	9.097	9795.3E6	6507.5E6	736.129	754.750
13) T	2,4-DB	8.881	9.605	1818.1E6	907.7E6	658.970	815.544
14) T	DINOSEB	9.940	9.934	8132.5E6	4883.3E6	749.026	762.854
15) T	Picloram	9.772	10.909	14051.1E6	8636.7E6	665.210	680.729
16) T	DCPA	10.207	10.865	13168.9E6	8235.4E6	727.669	771.257

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110822\
Data File : PS021082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Nov 2022 17:36
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 00:40:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
Quant Title : 8080.M
QLast Update : Mon Oct 31 10:07:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

