

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110918\
 Data File : PS002381.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2018 13:48
 Operator : EI\SJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Manual Integrations
 APPROVED

Sohil
 11/12/2018 1:43:22 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 15:48:21 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 15:03:49 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.427	10.920	291.6E6	360.8E6	951.229m	980.278
Target Compounds							
1) T	Dalapon	2.954	2.394	397.3E6	386.4E6	898.867	899.934
2) T	3,5-DICHL...	9.858	9.076	409.4E6	413.9E6	877.780	905.458
3) T	4-Nitroph...	11.021	10.067	146.9E6	163.5E6	906.415	928.411
5) T	DICAMBA	11.737	11.203	1281.3E6	1287.6E6	916.895	953.501m
6) T	MCPD	12.078	11.436	95820001	85846473	104.259m	95.366
7) T	MCPA	12.329	11.812	132.8E6	134.6E6	92.497m	95.960
8) T	DICHLORPROP	12.923	12.326	316.3E6	318.5E6	903.912	969.270m
9) T	2,4-D	13.280	12.797	305.8E6	364.2E6	959.631m	912.870
10) T	Pentachlo...	13.659	13.342	4397.8E6	4653.5E6	887.292	897.153
11) T	2,4,5-TP ...	14.527	13.960	1749.6E6	1900.2E6	919.763	935.100
12) T	2,4,5-T	14.923	14.523	1551.2E6	1497.5E6	920.188m	956.090
13) T	2,4-DB	15.670	15.175	185.3E6	210.0E6	964.449m	936.748
14) T	DINOSEB	17.148	15.602	1302.5E6	1382.2E6	899.661	917.181
15) T	Picloram	16.915	16.779	2100.7E6	2437.2E6	1004.742	906.919
16) T	DCPA	17.489	16.831	2168.9E6	2379.4E6	924.245m	1010.940

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110918\
Data File : PS002381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2018 13:48
Operator : EI\SJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Manual Integrations
APPROVED

Sohil
11/12/2018 1:43:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 15:48:21 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 15:03:49 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

