

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110918\
 Data File : PS002390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2018 18:54
 Operator : EI\SJ
 Sample : J5768-02MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 OK-02-110718-AMS

Manual Integrations
 APPROVED

Sohil
 11/12/2018 1:43:40 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 10 03:14:28 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 15:54:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.444	10.937	157.9E6	156.8E6	518.361	428.562
Target Compounds							
1) T	Dalapon	2.973	2.416	224.6E6	209.6E6	509.271	486.950
2) T	3,5-DICHL...	9.874	9.092	258.7E6	237.6E6	558.419	524.045m
3) T	4-Nitroph...	11.088	10.092	13306619	1925975	82.011	10.992 #
5) T	DICAMBA	11.753	11.219	746.0E6	698.7E6	538.413	516.460m
6) T	MCPD	12.093	11.450	46661308	55666823	49.717m	61.366
7) T	MCPA	12.342	11.825	84998283	84883435	58.849	60.320
8) T	DICHLORPROP	12.939	12.343	196.5E6	195.1E6	567.630	588.690
9) T	2,4-D	13.297	12.813	227.5E6	266.9E6	707.740	668.254
10) T	Pentachlo...	13.675	13.359	2543.6E6	2671.7E6	525.846	524.992
11) T	2,4,5-TP ...	14.543	13.977	1128.8E6	1243.1E6	600.685	616.258
12) T	2,4,5-T	14.940	14.542	1032.3E6	883.3E6	612.439	565.081
13) T	2,4-DB	15.686	15.191	129.3E6	142.9E6	667.536	630.358
14) T	DINOSEB	17.164	15.618	633.5E6	659.6E6	443.614	440.729
15) T	Picloram	16.931	16.796	1234.6E6	1114.7E6	581.354	419.013 #
16) T	DCPA	17.507	16.848	948.1E6	1430.8E6	409.265	596.607 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110918\
Data File : PS002390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2018 18:54
Operator : EI\SJ
Sample : J5768-02MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
OK-02-110718-AMS

Manual Integrations
APPROVED

Sohil
11/12/2018 1:43:40 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 10 03:14:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 15:54:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

