

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110920\
 Data File : PS013161.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2020 16:26
 Operator : DD\AJ
 Sample : PB132900BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB132900BS

Manual Integrations
 APPROVED

Ankita
 11/10/2020 8:39:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 17:24:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	6.913	457.1E6	217.2E6	511.298	558.340
Target Compounds							
1) T	Dalapon	3.030	2.623	558.9E6	346.9E6	528.244	654.605m
2) T	3,5-DICHL...	6.426	6.066	668.3E6	308.5E6	542.664	501.439
3) T	4-Nitroph...	7.045	6.526	213.2E6	99787959	499.838	419.015
5) T	DICAMBA	7.399	7.078	1759.0E6	846.6E6	492.533	509.971
6) T	MCPD	7.565	7.171	152.1E6	69650730	57.125	52.622
7) T	MCPA	7.714	7.384	194.5E6	112.4E6	50.926	52.822m
8) T	DICHLORPROP	8.090	7.704	454.2E6	237.1E6	507.153	530.544
9) T	2,4-D	8.322	8.001	519.5E6	308.2E6	549.883	580.713
10) T	Pentachlo...	8.629	8.425	7050.6E6	3123.9E6	547.824	565.277
11) T	2,4,5-TP ...	9.177	8.785	2779.3E6	1291.5E6	541.197	557.291
12) T	2,4,5-T	9.464	9.168	2717.3E6	1218.7E6	513.456	550.200
13) T	2,4-DB	10.018	9.674	290.6E6	158.3E6	344.921	487.777 #
14) T	DINOSEB	11.171	10.002	2044.0E6	884.7E6	495.702	524.742
15) T	Picloram	11.008	10.999	2519.2E6	1474.2E6	383.110	408.432
16) T	DCPA	11.437	10.928	3155.6E6	1654.3E6	485.913	551.004

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110920\
Data File : PS013161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2020 16:26
Operator : DD\AJ
Sample : PB132900BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB132900BS

Manual Integrations
APPROVED

Ankita
11/10/2020 8:39:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 17:24:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

