

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110924\
 Data File : PS028385.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2024 02:58
 Operator : AR\AJ
 Sample : P4602-45MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

CC0R8MSD

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 11/11/2024

Supervised By :Ankita Jodhani 11/12/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 03:12:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 08 17:46:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.244	7.752	1092.5E6	468.4E6	395.733	293.930 #
Target Compounds							
1) T	Dalapon	2.644	2.706	510.3E6	629.9E6	143.028	222.133m#
2) T	3,5-DICHL...	6.415	6.704	733.4E6	392.1E6	183.282	165.524
3) T	4-Nitroph...	7.046	7.279	309.7E6	251.0E6	167.035	236.978 #
5) T	DICAMBA	7.434	7.953	2853.8E6	1676.6E6	237.011	233.853
6) T	MCPD	7.613	8.049	78809431	96677123	9.636	16.457 #
7) T	MCPA	7.764	8.296	256.2E6	200.0E6	22.849	24.970
8) T	DICHLORPROP	8.143	8.670	494.0E6	261.9E6	159.363	147.006
9) T	2,4-D	8.375	9.000	1069.6E6	469.0E6	304.366	242.547
10) T	Pentachlo...	8.674	9.531	2657.5E6	1357.9E6	59.000	54.496
11) T	2,4,5-TP ...	9.254	9.907	1572.9E6	974.6E6	83.097m	92.016
12) T	2,4,5-T	9.547	10.328	2737.3E6	1613.8E6	141.237	155.528
13) T	2,4-DB	10.124	10.893	294.0E6	203.8E6	95.112	153.812 #
14) T	DINOSEB	11.353	11.279	627.4E6	255.0E6	38.695	35.523
15) T	Picloram	11.145	12.374	10000.9E6	4078.8E6	318.440	283.396
16) T	DCPA	11.630	12.321	9819.5E6	4191.3E6	367.407	373.882

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS110924\
Data File : PS028385.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2024 02:58
Operator : AR\AJ
Sample : P4602-45MSD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
CC0R8MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/11/2024
Supervised By :Ankita Jodhani 11/12/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 03:12:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
Quant Title : 8080.M
QLast Update : Fri Nov 08 17:46:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

