

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111025\
 Data File : PS032330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Nov 2025 15:29
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/11/2025
 Supervised By :Yogesh Patel 11/11/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 01:49:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.283	7.616	3469.2E6	2832.0E6	765.607	776.835
Target Compounds							
1) T	Dalapon	2.655	2.602	4743.6E6	4650.3E6	660.784	668.724
2) T	3,5-DICHL...	6.449	6.584	4661.1E6	4381.1E6	720.838	729.102
3) T	4-Nitroph...	7.103	7.184	970.8E6	1737.3E6	723.125	765.343
5) T	DICAMBA	7.465	7.806	13795.3E6	12784.9E6	731.219	740.191
6) T	MCPD	7.642	7.904	811.2E6	1008.8E6	74.493	75.844
7) T	MCPA	7.791	8.149	983.2E6	1176.7E6	73.231	72.169
8) T	DICHLORPROP	8.173	8.521	3152.2E6	3482.4E6	708.566	730.198
9) T	2,4-D	8.413	8.863	3238.3E6	3863.0E6	717.749	736.419
10) T	Pentachlo...	8.707	9.356	47534.1E6	44506.9E6	757.646	758.175
11) T	2,4,5-TP ...	9.278	9.744	19179.1E6	19865.2E6	737.332	739.750
12) T	2,4,5-T	9.578	10.171	18058.0E6	17799.5E6	757.524	750.461
13) T	2,4-DB	10.155	10.737	2005.7E6	2799.3E6	761.132	742.269
14) T	DINOSEB	11.345	11.104	15004.7E6	12836.0E6	721.116	725.717
15) T	Picloram	11.184	12.225	15351.0E6	25414.2E6	723.231m	683.505
16) T	DCPA	11.645	12.139	24917.9E6	23569.6E6	742.665	739.774

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111025\
Data File : PS032330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Nov 2025 15:29
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/11/2025
Supervised By :Yogesh Patel 11/11/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 01:49:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

