

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111119\
 Data File : PS007648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2019 17:24
 Operator : SM\AJ
 Sample : PB124673BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB124673BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 11 18:06:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 06 02:45:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.444	5.167	577.9E6	320.0E6	419.609	455.478
Target Compounds							
1) T	Dalapon	2.005	1.712	110.1E6	171.8E6	60.876	144.646 #
2) T	3,5-DICHL...	0.000	4.502	0	61099637	N.D.	67.631
5) T	DICAMBA	5.534	0.000	41704451	0	8.198	N.D. #
7) T	MCPA	5.910	5.536	37291347	32065182	6.775	10.668 #
9) T	2,4-D	0.000	6.006	0	22322547	N.D.	27.682
10) T	Pentachlo...	0.000	6.378	0	27294836	N.D.	2.787
11) T	2,4,5-TP ...	0.000	6.764	0	36142185	N.D.	9.230
12) T	2,4,5-T	7.411	0.000	121.2E6	0	14.806	N.D. #
15) T	Picloram	8.982	8.918	108.5E6	99100211	9.774	16.914 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111119\
Data File : PS007648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2019 17:24
Operator : SM\AJ
Sample : PB124673BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB124673BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 11 18:06:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
Quant Title : 8080.M
QLast Update : Wed Nov 06 02:45:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.005 min
Delta R.T.: -0.007 min
Response: 110135843
Conc: 60.88 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB124673BL




#1 Dalapon

R.T.: 1.712 min
Delta R.T.: -0.011 min
Response: 171752770
Conc: 144.65 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.804 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.502 min
Delta R.T.: 0.040 min
Response: 61099637
Conc: 67.63 ng/ml



#4 2,4-DCAA


R.T.: 5.444 min
Delta R.T.: 0.015 min
Response: 577879008
Conc: 419.61 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB124673BL




#4 2,4-DCAA

R.T.: 5.167 min
Delta R.T.: 0.015 min
Response: 320002457
Conc: 455.48 ng/ml




#5 DICAMBA

R.T.: 5.534 min
Delta R.T.: -0.043 min
Response: 41704451
Conc: 8.20 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 5.277 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 5.910 min
Delta R.T.: 0.050 min
Response: 37291347
Conc: 6.77 ug/ml

Instrument :
ECD_S
ClientSampleId :
PB124673BL




#7 MCPA

R.T.: 5.536 min
Delta R.T.: 0.000 min
Response: 32065182
Conc: 10.67 ug/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 6.414 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 6.006 min
Delta R.T.: -0.066 min
Response: 22322547
Conc: 27.68 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 6.671 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
PB124673BL




#10 Pentachlorophenol

R.T.: 6.378 min
Delta R.T.: -0.057 min
Response: 27294836
Conc: 2.79 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 7.203 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 6.764 min
Delta R.T.: -0.033 min
Response: 36142185
Conc: 9.23 ng/ml



#12 2,4,5-T


R.T.: 7.411 min
Delta R.T.: -0.064 min
Response: 121203803
Conc: 14.81 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB124673BL




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 7.156 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 8.982 min
Delta R.T.: 0.064 min
Response: 108527679
Conc: 9.77 ng/ml



#15 Picloram

R.T.: 8.918 min
Delta R.T.: 0.036 min
Response: 99100211
Conc: 16.91 ng/ml