

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111122\
 Data File : PS021130.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2022 10:11
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 09:47:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 31 10:07:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.410	6.843	1640.7E6	1140.1E6	786.112	824.281
Target Compounds							
1) T	Dalapon	2.506	2.574	1870.3E6	1327.0E6	706.696	729.611
2) T	3,5-DICHL...	5.778	6.016	2200.6E6	1603.9E6	718.613	785.881
3) T	4-Nitroph...	6.248	6.453	1015.0E6	640.6E6	674.836	745.038
5) T	DICAMBA	6.556	7.010	6540.3E6	4850.7E6	740.294	781.101
6) T	MCPD	6.715	7.111	488.5E6	374.1E6	74.850	78.743
7) T	MCPA	6.842	7.323	677.9E6	511.1E6	72.245	75.327
8) T	DICHLORPROP	7.155	7.636	1660.9E6	1192.0E6	727.618	774.119
9) T	2,4-D	7.356	7.927	2037.4E6	1456.6E6	729.022	777.737
10) T	Pentachlo...	7.590	8.358	23379.3E6	16554.0E6	741.985	780.413
11) T	2,4,5-TP ...	8.117	8.720	9798.9E6	6976.3E6	737.605	777.399
12) T	2,4,5-T	8.375	9.096	9887.7E6	6710.1E6	743.068	778.247
13) T	2,4-DB	8.885	9.604	2016.7E6	888.6E6	730.945	798.366
14) T	DINOSEB	9.945	9.934	8077.1E6	4979.8E6	743.922	777.927
15) T	Picloram	9.775	10.906	15246.6E6	9438.0E6	721.807	743.881
16) T	DCPA	10.211	10.864	13247.4E6	8361.0E6	732.011	783.019

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111122\
Data File : PS021130.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2022 10:11
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 09:47:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
Quant Title : 8080.M
QLast Update : Mon Oct 31 10:07:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

