

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111122\
 Data File : PS021141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2022 15:51
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 09:49:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 31 10:07:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.408	6.844	1700.0E6	1187.4E6	814.538	858.479
Target Compounds							
1) T	Dalapon	2.506	2.577	1915.6E6	1373.8E6	723.820	755.354
2) T	3,5-DICHL...	5.776	6.016	2285.4E6	1681.7E6	746.314	823.999
3) T	4-Nitroph...	6.245	6.454	1051.1E6	669.0E6	698.826	778.117
5) T	DICAMBA	6.553	7.011	6741.8E6	5031.7E6	763.109	810.242
6) T	MCPD	6.712	7.111	515.5E6	383.4E6	78.992	80.704
7) T	MCPA	6.839	7.324	702.9E6	539.2E6	74.907	79.461
8) T	DICHLORPROP	7.152	7.636	1723.7E6	1248.7E6	755.136	810.914
9) T	2,4-D	7.353	7.926	2084.8E6	1507.8E6	745.953	805.100
10) T	Pentachlo...	7.586	8.359	24008.7E6	17128.6E6	761.960	807.499
11) T	2,4,5-TP ...	8.114	8.720	10139.2E6	7263.2E6	763.223	809.372
12) T	2,4,5-T	8.372	9.096	10188.1E6	6964.6E6	765.641	807.761
13) T	2,4-DB	8.880	9.603	2069.3E6	932.4E6	750.034	837.744
14) T	DINOSEB	9.940	9.933	8382.2E6	5159.6E6	772.024	806.021m
15) T	Picloram	9.770	10.905	15226.0E6	9469.2E6	720.834	746.345
16) T	DCPA	10.207	10.863	13726.1E6	8726.4E6	758.459	817.238

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111122\
Data File : PS021141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2022 15:51
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 09:49:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
Quant Title : 8080.M
QLast Update : Mon Oct 31 10:07:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

