

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111125\
 Data File : PS032353.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2025 22:09
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/12/2025
 Supervised By :Yogesh Patel 11/12/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 01:17:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.280	7.614	3454.3E6	2952.2E6	762.326m	809.822m
Target Compounds							
1) T	Dalapon	2.655	2.603	4591.4E6	4546.8E6	639.581	653.841
2) T	3,5-DICHL...	6.447	6.583	4753.9E6	4425.8E6	735.197	736.533
3) T	4-Nitroph...	7.098	7.178	1028.3E6	1754.7E6	765.967	773.007
5) T	DICAMBA	7.462	7.806	13970.9E6	12915.8E6	740.527	747.770
6) T	MCPD	7.640	7.903	805.5E6	1030.4E6	73.968	77.465
7) T	MCPA	7.789	8.148	1031.7E6	1257.5E6	76.845	77.124m
8) T	DICHLORPROP	8.170	8.519	3183.4E6	3532.5E6	715.572	740.699
9) T	2,4-D	8.409	8.860	3332.4E6	3899.1E6	738.603	743.298
10) T	Pentachlo...	8.700	9.355	47707.8E6	44591.4E6	760.415	759.614
11) T	2,4,5-TP ...	9.275	9.742	19430.9E6	19924.1E6	747.012	741.944
12) T	2,4,5-T	9.573	10.170	18524.6E6	18160.1E6	777.096	765.666
13) T	2,4-DB	10.150	10.735	2127.1E6	2784.2E6	807.212	738.257
14) T	DINOSEB	11.340	11.103	14995.7E6	12879.7E6	720.686	728.183
15) T	Picloram	11.178	12.221	16053.7E6	23139.5E6	756.338	622.330
16) T	DCPA	11.640	12.136	25296.3E6	23474.0E6	753.942	736.773m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111125\
Data File : PS032353.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2025 22:09
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/12/2025
Supervised By :Yogesh Patel 11/12/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 01:17:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

