

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111125\
 Data File : PS032366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2025 04:09
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/12/2025
 Supervised By :Yogesh Patel 11/12/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 04:37:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.280	7.614	3288.8E6	2932.8E6	725.804m	804.488
Target Compounds							
1) T	Dalapon	2.654	2.603	4564.6E6	4537.5E6	635.845	652.500
2) T	3,5-DICHL...	6.447	6.582	4783.7E6	4445.8E6	739.794	739.869
3) T	4-Nitroph...	7.097	7.176	1047.4E6	1774.0E6	780.195	781.526
5) T	DICAMBA	7.462	7.806	14039.2E6	12970.3E6	744.149	750.927
6) T	MCPD	7.640	7.904	772.3E6	1032.2E6	70.922	77.607
7) T	MCPA	7.790	8.148	1035.9E6	1232.8E6	77.159	75.609
8) T	DICHLORPROP	8.170	8.519	3200.0E6	3549.1E6	719.313	744.182
9) T	2,4-D	8.408	8.860	3367.2E6	3930.4E6	746.315	749.270
10) T	Pentachlo...	8.701	9.356	47935.5E6	44708.1E6	764.044	761.603
11) T	2,4,5-TP ...	9.275	9.742	19569.9E6	19961.1E6	752.352	743.322
12) T	2,4,5-T	9.573	10.170	18310.5E6	18311.8E6	768.114	772.061
13) T	2,4-DB	10.149	10.735	2151.2E6	2815.5E6	816.360	746.556
14) T	DINOSEB	11.341	11.104	15040.7E6	12917.9E6	722.846	730.344
15) T	Picloram	11.178	12.222	15481.9E6	21936.3E6	729.397	589.969
16) T	DCPA	11.641	12.138	25422.7E6	23594.1E6	757.710	740.542

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111125\
Data File : PS032366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2025 04:09
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/12/2025
Supervised By :Yogesh Patel 11/12/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 04:37:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

