

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111218\
 Data File : PS002419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2018 23:07
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 11/13/2018 4:35:03 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 06:40:58 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 15:54:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.438	10.932	234.0E6	269.7E6	768.171m	737.156
Target Compounds							
1) T	Dalapon	2.965	2.406	298.7E6	292.0E6	677.352	678.534
2) T	3,5-DICHL...	9.870	9.087	323.4E6	310.1E6	698.070	684.012
3) T	4-Nitroph...	11.034	10.079	113.3E6	118.2E6	698.432	674.770
5) T	DICAMBA	11.747	11.214	973.7E6	964.1E6	702.684	712.672
6) T	MCPD	12.087	11.445	68075667	64600635	72.534	71.215
7) T	MCPA	12.337	11.821	104.8E6	99599811	72.580m	70.778
8) T	DICHLORPROP	12.934	12.338	239.7E6	240.5E6	692.319	725.665
9) T	2,4-D	13.294	12.809	237.5E6	280.7E6	738.978m	702.867
10) T	Pentachlo...	13.670	13.353	3452.5E6	3666.1E6	713.743	720.389
11) T	2,4,5-TP ...	14.539	13.972	1335.7E6	1437.4E6	710.807	712.574
12) T	2,4,5-T	14.936	14.537	1207.9E6	1067.0E6	716.618	682.577
13) T	2,4-DB	15.683	15.187	147.5E6	162.5E6	761.312m	716.690
14) T	DINOSEB	17.159	15.614	991.3E6	1040.4E6	694.164	695.134
15) T	Picloram	16.928	16.790	1393.5E6	1917.1E6	656.159	720.631
16) T	DCPA	17.501	16.844	1674.7E6	1625.7E6	722.925	677.852

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111218\
Data File : PS002419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2018 23:07
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
11/13/2018 4:35:03 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 06:40:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 15:54:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

