

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111221\
 Data File : PS017390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2021 15:30
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 22:04:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.150	6.738	1068.4E6	418.0E6	699.018	715.162
Target Compounds							
1) T	Dalapon	2.987	2.492	1509.2E6	445.1E6	695.023	687.843
2) T	3,5-DICHL...	6.374	5.936	1441.4E6	591.2E6	667.739	680.681
3) T	4-Nitroph...	6.982	6.365	513.7E6	204.5E6	589.056	634.259
5) T	DICAMBA	7.338	6.898	4123.7E6	1788.2E6	688.466	702.352
6) T	MCPD	7.507	6.994	306.7E6	142.8E6	68.611	70.925
7) T	MCPA	7.655	7.205	430.9E6	196.6E6	69.848	66.402
8) T	DICHLORPROP	8.029	7.515	1024.2E6	445.8E6	683.108	693.211
9) T	2,4-D	8.259	7.807	1213.8E6	509.1E6	685.283	676.420
10) T	Pentachlo...	8.562	8.229	14549.9E6	5845.3E6	701.265	715.930
11) T	2,4,5-TP ...	9.113	8.590	5679.6E6	2453.9E6	698.370	711.861
12) T	2,4,5-T	9.398	8.967	5668.4E6	2245.3E6	705.717	708.402
13) T	2,4-DB	9.948	9.471	1009.4E6	271.7E6	670.971	655.070
14) T	DINOSEB	11.096	9.795	4339.0E6	1713.7E6	765.964	776.635
15) T	Picloram	10.921	10.775	7617.2E6	2717.6E6	713.124	709.612
16) T	DCPA	11.364	10.724	7340.2E6	3061.7E6	720.741	717.485

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111221\
Data File : PS017390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2021 15:30
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 22:04:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

