

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111221\
 Data File : PS017402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2021 22:01
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 22:35:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
 Quant Title : 8080.M
 QLast Update : Tue Oct 19 01:21:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.150	6.738	1078.2E6	418.3E6	705.398	715.768
Target Compounds							
1) T	Dalapon	2.987	2.492	1472.5E6	433.9E6	678.114	670.659
2) T	3,5-DICHL...	6.374	5.936	1448.5E6	589.4E6	671.010	678.599
3) T	4-Nitroph...	6.981	6.365	539.2E6	210.0E6	618.201	651.569
5) T	DICAMBA	7.338	6.898	4135.2E6	1775.9E6	690.384	697.497
6) T	MCPD	7.506	6.994	308.2E6	141.7E6	68.948	70.381
7) T	MCPA	7.655	7.205	435.7E6	199.7E6	70.619	67.454
8) T	DICHLORPROP	8.029	7.515	1028.9E6	443.8E6	686.214	690.049
9) T	2,4-D	8.260	7.806	1240.1E6	514.4E6	700.150	683.563
10) T	Pentachlo...	8.562	8.229	14566.0E6	5785.5E6	702.042	708.601
11) T	2,4,5-TP ...	9.112	8.590	5694.7E6	2434.9E6	700.226	706.359
12) T	2,4,5-T	9.397	8.967	5740.6E6	2248.8E6	714.701	709.511
13) T	2,4-DB	9.947	9.471	1072.0E6	276.5E6	712.597	666.642
14) T	DINOSEB	11.096	9.796	4342.7E6	1692.0E6	766.604	766.778
15) T	Picloram	10.922	10.776	7231.8E6	2547.7E6	677.043	665.248
16) T	DCPA	11.364	10.725	7364.3E6	3037.5E6	723.110	711.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111221\
Data File : PS017402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2021 22:01
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 22:35:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101821.M
Quant Title : 8080.M
QLast Update : Tue Oct 19 01:21:46 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

