

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111224\
 Data File : PS028398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 15:14
 Operator : AR\AJ
 Sample : PB164793TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 CC0R3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 23:51:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 08 17:46:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.244	7.746	975.4E6	453.8E6	353.292	284.803
Target Compounds							
1) T	Dalapon	2.600f	2.668f	60093564	298.5E6	16.842	105.251 #
2) T	3,5-DICHL...	6.401	6.728	9229271	14025739	2.307	5.920 #
3) T	4-Nitroph...	0.000	7.312f	0	16121459	N.D.	15.223
5) T	DICAMBA	7.470f	7.934	-2810321	9646174	N.D.	1.345
6) T	MCPD	7.668f	8.042	91304053	7772360	11.164	1.323 #
7) T	MCPA	7.765	8.271f	18740424	15290988	1.671	1.910
8) T	DICHLORPROP	8.160	8.715f	14426497	57465175	4.654	32.261 #
9) T	2,4-D	8.372	8.991	979411	4144991	<MDL	2.143 #
10) T	Pentachlo...	8.671	9.524	7866569	6870167	<MDL	<MDL #
11) T	2,4,5-TP ...	9.271	9.953f	-1810928	27442725	N.D.	2.591
12) T	2,4,5-T	0.000	10.296f	0	40590510	N.D.	3.912
13) T	2,4-DB	10.111	10.924	51896312	2236500	16.790	1.688 #
14) T	DINOSEB	11.328	11.274	36352523	7110009	2.242	<MDL #
15) T	Picloram	11.135	12.370	68016118	5265831	2.166	<MDL #
16) T	DCPA	11.603f	12.322	27561236	1242082	1.031	<MDL #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111224\
Data File : PS028398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 15:14
Operator : AR\AJ
Sample : PB164793TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
CC0R3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 23:51:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
Quant Title : 8080.M
QLast Update : Fri Nov 08 17:46:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.600 min
Delta R.T.: -0.050 min
Response: 60093564
Conc: 16.84 ng/ml

Instrument :
ECD_S
ClientSampleId :
CC0R3




#1 Dalapon

R.T.: 2.668 min
Delta R.T.: -0.039 min
Response: 298456788
Conc: 105.25 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.401 min
Delta R.T.: -0.019 min
Response: 9229271
Conc: 2.31 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.728 min
Delta R.T.: 0.021 min
Response: 14025739
Conc: 5.92 ng/ml



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T.: 7.049 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
CC0R3




#3 4-Nitrophenol

R.T.: 7.312 min
Delta R.T.: 0.030 min
Response: 16121459
Conc: 15.22 ng/ml




#4 2,4-DCAA

R.T.: 7.244 min
Delta R.T.: -0.006 min
Response: 975351053
Conc: 353.29 ng/ml



#4 2,4-DCAA


R.T.: 7.746 min
Delta R.T.: -0.009 min
Response: 453811757
Conc: 284.80 ng/ml



#5 DICAMBA


R.T.: 7.470 min
Delta R.T.: 0.031 min
Response: -2810321
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
CC0R3




#5 DICAMBA

R.T.: 7.934 min
Delta R.T.: -0.023 min
Response: 9646174
Conc: 1.35 ng/ml




#6 MCP

R.T.: 7.668 min
Delta R.T.: 0.040 min
Response: 91304053
Conc: 11.16 ug/ml



#6 MCP


R.T.: 8.042 min
Delta R.T.: -0.023 min
Response: 7772360
Conc: 1.32 ug/ml



#7 MCPA


R.T.: 7.765 min
Delta R.T.: -0.014 min
Response: 18740424
Conc: 1.67 ug/ml

Instrument :
ECD_S
ClientSampleId :
CC0R3




#7 MCPA

R.T.: 8.271 min
Delta R.T.: -0.041 min
Response: 15290988
Conc: 1.91 ug/ml




#8 DICHLORPROP

R.T.: 8.160 min
Delta R.T.: 0.011 min
Response: 14426497
Conc: 4.65 ng/ml



#8 DICHLORPROP


R.T.: 8.715 min
Delta R.T.: 0.041 min
Response: 57465175
Conc: 32.26 ng/ml



#9 2,4-D


R.T.: 8.372 min
Delta R.T.: -0.007 min
Response: 979411
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
CC0R3




#9 2,4-D

R.T.: 8.991 min
Delta R.T.: -0.014 min
Response: 4144991
Conc: 2.14 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.271 min
Delta R.T.: 0.010 min
Response: -1810928
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 9.953 min
Delta R.T.: 0.041 min
Response: 27442725
Conc: 2.59 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 9.553 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
CC0R3




#12 2,4,5-T

R.T.: 10.296 min
Delta R.T.: -0.038 min
Response: 40590510
Conc: 3.91 ng/ml




#13 2,4-DB

R.T.: 10.111 min
Delta R.T.: -0.019 min
Response: 51896312
Conc: 16.79 ng/ml



#13 2,4-DB


R.T.: 10.924 min
Delta R.T.: 0.024 min
Response: 2236500
Conc: 1.69 ng/ml



#14 DINoseb


R.T.: 11.328 min
Delta R.T.: -0.017 min
Response: 36352523
Conc: 2.24 ng/ml

Instrument :
ECD_S
ClientSampleId :
CC0R3




#14 DINoseb

R.T.: 11.274 min
Delta R.T.: -0.007 min
Response: 7110009
Conc: N.D.




#15 Picloram

R.T.: 11.135 min
Delta R.T.: -0.015 min
Response: 68016118
Conc: 2.17 ng/ml



#15 Picloram


R.T.: 12.370 min
Delta R.T.: -0.009 min
Response: 5265831
Conc: N.D.



#16 DCPA

R.T.: 11.603 min
Delta R.T.: -0.034 min
Response: 27561236
Conc: 1.03 ng/ml

Instrument :
ECD_S
ClientSampleId :
CC0R3



#16 DCPA

R.T.: 12.322 min
Delta R.T.: -0.004 min
Response: 1242082
Conc: N.D.