

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111320\
 Data File : PS013201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2020 20:06
 Operator : DD\AJ
 Sample : L4681-03MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-1-SMSD

Manual Integrations
 APPROVED

Ankita
 11/25/2020 8:33:21 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 03:19:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.208	6.912	429.3E6	191.6E6	480.271	492.327m
Target Compounds							
1) T	Dalapon	3.029	2.621	283.0E6	196.7E6	267.439	371.139m#
2) T	3,5-DICHL...	6.426	6.067	356.9E6	165.8E6	289.791	269.451m
3) T	4-Nitroph...	7.048	6.535	28890541	37615666	67.741	157.950m#
5) T	DICAMBA	7.398	7.078	1028.7E6	487.9E6	288.063	293.912
6) T	MCPD	7.563	7.169	79569182	36218354	29.890	27.364
7) T	MCPA	7.709	7.382	184.3E6	91180043	48.276	42.855
8) T	DICHLORPROP	8.090	7.703	246.7E6	134.4E6	275.533	300.635m
9) T	2,4-D	8.323	8.001	296.7E6	185.8E6	314.096	350.058m
10) T	Pentachlo...	8.628	8.425	3523.1E6	1597.7E6	273.744	289.119
11) T	2,4,5-TP ...	9.177	8.786	1473.9E6	672.9E6	286.999	290.374m
12) T	2,4,5-T	9.464	9.168	1472.6E6	681.6E6	278.259	307.722
13) T	2,4-DB	10.017	9.675	152.5E6	80704795	181.007	248.671m#
14) T	DINOSEB	11.170	10.003	192.4E6	78571385	46.648	46.605m
15) T	Picloram	11.000	10.992	1821.8E6	1018.3E6	277.056	282.114
16) T	DCPA	11.438	10.929	1845.1E6	1033.7E6	284.115	344.318

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111320\
Data File : PS013201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2020 20:06
Operator : DD\AJ
Sample : L4681-03MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TP-1-SMSD

Manual Integrations
APPROVED

Ankita
11/25/2020 8:33:21 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 03:19:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

