

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111420\
 Data File : PS013233.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Nov 2020 19:02
 Operator : DD\AJ
 Sample : L4719-12
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-3A-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 11:31:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.211	6.913	318.4E6	165.9E6	356.227	426.476
Target Compounds							
1) T	Dalapon	2.992f	2.590	67558844	115.2E6	63.855	217.418 #
2) T	3,5-DICHL...	0.000	6.095	0	84313325	N.D.	137.056
3) T	4-Nitroph...	7.098f	0.000	6639918	0	15.569	N.D. #
5) T	DICAMBA	7.386	7.046	12754534	7048434	3.571	4.246
7) T	MCPA	7.705	0.000	168.4E6	0	44.092	N.D. #
8) T	DICHLORPROP	0.000	7.745	0	8545355	N.D.	19.119
9) T	2,4-D	0.000	7.972	0	77776	N.D.	<MDL
10) T	Pentachlo...	8.683f	0.000	10516813	0	<MDL	N.D. #
11) T	2,4,5-TP ...	0.000	8.742f	0	49209517	N.D.	21.234
12) T	2,4,5-T	9.533f	0.000	13486630	0	2.548	N.D. #
14) T	DINOSEB	11.143	10.033	20674587	7342056	5.014	4.355
15) T	Picloram	10.975	11.031	34583358	-496096	5.259	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111420\
Data File : PS013233.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Nov 2020 19:02
Operator : DD\AJ
Sample : L4719-12
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-3A-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 11:31:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.992 min
Delta R.T.: -0.040 min
Response: 67558844
Conc: 63.85 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3A-S




#1 Dalapon

R.T.: 2.590 min
Delta R.T.: -0.033 min
Response: 115226411
Conc: 217.42 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.427 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.095 min
Delta R.T.: 0.028 min
Response: 84313325
Conc: 137.06 ng/ml



#3 4-Nitrophenol


R.T.: 7.098 min
Delta R.T.: 0.053 min
Response: 6639918
Conc: 15.57 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3A-S




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 6.527 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.211 min
Delta R.T.: 0.000 min
Response: 318441675
Conc: 356.23 ng/ml



#4 2,4-DCAA


R.T.: 6.913 min
Delta R.T.: 0.000 min
Response: 165937399
Conc: 426.48 ng/ml



#5 DICAMBA


R.T.: 7.386 min
Delta R.T.: -0.014 min
Response: 12754534
Conc: 3.57 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3A-S




#5 DICAMBA

R.T.: 7.046 min
Delta R.T.: -0.034 min
Response: 7048434
Conc: 4.25 ng/ml




#7 MCPA

R.T.: 7.705 min
Delta R.T.: -0.011 min
Response: 168361187
Conc: 44.09 ug/ml



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 7.389 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.092 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
TP-3A-S




#8 DICHLORPROP

R.T.: 7.745 min
Delta R.T.: 0.040 min
Response: 8545355
Conc: 19.12 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 9.179 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 8.742 min
Delta R.T.: -0.047 min
Response: 49209517
Conc: 21.23 ng/ml



#12 2,4,5-T


R.T.: 9.533 min
Delta R.T.: 0.066 min
Response: 13486630
Conc: 2.55 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3A-S




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.170 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 11.143 min
Delta R.T.: -0.029 min
Response: 20674587
Conc: 5.01 ng/ml



#14 DINOSEB


R.T.: 10.033 min
Delta R.T.: 0.028 min
Response: 7342056
Conc: 4.35 ng/ml



#15 Picloram

R.T.: 10.975 min
Delta R.T.: -0.031 min
Response: 34583358
Conc: 5.26 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3A-S



#15 Picloram

R.T.: 11.031 min
Delta R.T.: 0.032 min
Response: -496096
Conc: N.D.