

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111422\
 Data File : PS021152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2022 10:30
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 14 13:43:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 31 10:07:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.408	6.844	1594.9E6	1092.2E6	764.179	789.619
Target Compounds							
1) T	Dalapon	2.504	2.576	1801.9E6	1276.0E6	680.859	701.565
2) T	3,5-DICHL...	5.776	6.016	2137.0E6	1548.2E6	697.850	758.605
3) T	4-Nitroph...	6.246	6.454	965.0E6	610.0E6	641.603	709.433
5) T	DICAMBA	6.552	7.011	6380.6E6	4700.8E6	722.223	756.956
6) T	MCPD	6.711	7.111	472.9E6	355.1E6	72.452	74.740
7) T	MCPA	6.838	7.323	659.6E6	498.1E6	70.296	73.417
8) T	DICHLORPROP	7.152	7.636	1614.4E6	1154.9E6	707.236	750.052
9) T	2,4-D	7.352	7.927	1947.6E6	1387.8E6	696.877	740.983
10) T	Pentachlo...	7.586	8.358	22896.1E6	16137.1E6	726.650	760.757
11) T	2,4,5-TP ...	8.113	8.719	9483.4E6	6776.9E6	713.857	755.189
12) T	2,4,5-T	8.371	9.095	9612.1E6	6434.3E6	722.361	746.255
13) T	2,4-DB	8.880	9.603	1706.1E6	837.9E6	618.390	752.791
14) T	DINOSEB	9.939	9.933	7933.2E6	4765.7E6	730.673	744.480
15) T	Picloram	9.771	10.905	14067.1E6	8668.7E6	665.968	683.247
16) T	DCPA	10.205	10.862	12946.6E6	8140.4E6	715.385	762.361

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111422\
Data File : PS021152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2022 10:30
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 14 13:43:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102822.M
Quant Title : 8080.M
QLast Update : Mon Oct 31 10:07:57 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

