

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111424\
 Data File : PS028540.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Nov 2024 18:38
 Operator : AR\AJ
 Sample : P4821-05MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BP-F-2MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/15/2024
 Supervised By :Ankita Jodhani 11/15/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 15 00:24:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 13 13:41:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.240	7.739	1195.6E6	543.8E6	467.799	420.794
Target Compounds							
1) T	Dalapon	2.631	2.698	687.2E6	651.7E6	204.571	269.307m#
2) T	3,5-DICHL...	6.410	6.693	1417.8E6	677.2E6	379.963	350.442
3) T	4-Nitroph...	7.038	7.265	578.7E6	306.7E6	340.747m	355.249
5) T	DICAMBA	7.428	7.940	4169.6E6	2132.4E6	381.062	373.842
6) T	MCP	7.609	8.040	223.5E6	139.9E6	31.774	32.523
7) T	MCPA	7.759	8.284	349.8E6	216.6E6	34.275	34.971
8) T	DICHLORPROP	8.138	8.656	1078.4E6	508.5E6	365.400	348.137
9) T	2,4-D	8.368	8.987	1357.8E6	650.0E6	424.735	415.162
10) T	Pentachlo...	8.668	9.516	13093.1E6	5922.6E6	313.711	295.461
11) T	2,4,5-TP ...	9.249	9.893	6072.4E6	2948.2E6	349.578	349.388
12) T	2,4,5-T	9.542	10.314	6333.8E6	2929.2E6	356.339	349.454
13) T	2,4-DB	10.118	10.881	856.0E6	334.0E6	314.047	318.764
14) T	DINOSEB	11.331	11.261	666.4E6	201.6E6	46.741	36.460
15) T	Picloram	11.138	12.357	8670.4E6	3226.5E6	307.899	285.048
16) T	DCPA	11.623	12.305	8419.8E6	3838.0E6	319.983	395.737

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111424\
Data File : PS028540.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Nov 2024 18:38
Operator : AR\AJ
Sample : P4821-05MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BP-F-2MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/15/2024
Supervised By :Ankita Jodhani 11/15/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 15 00:24:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
Quant Title : 8080.M
QLast Update : Wed Nov 13 13:41:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

