

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111822\
 Data File : PS021198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Nov 2022 13:08
 Operator : AR\AJ
 Sample : N5598-02MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FRAC-TANK-0760533MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 00:32:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111622.M
 Quant Title : 8080.M
 QLast Update : Thu Nov 17 00:46:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.406	6.842	814.8E6	578.6E6	374.765	382.641
Target Compounds							
1) T	Dalapon	2.501	2.574	844.5E6	915.0E6	303.448m	453.952 #
2) T	3,5-DICHL...	5.774	6.015	1382.6E6	996.2E6	437.282	428.625
5) T	DICAMBA	6.550	7.008	4015.1E6	2947.9E6	437.738	428.279
6) T	MCP	6.706	7.104	300.6E6	221.9E6	44.499	42.555
7) T	MCPA	6.831	7.316	406.2E6	306.4E6	42.263	41.541
8) T	DICHLORPROP	7.150	7.633	1053.6E6	810.5E6	447.972	474.714
9) T	2,4-D	7.351	7.925	1326.6E6	993.2E6	467.651	488.386
10) T	Pentachlo...	7.582	8.355	13822.6E6	9548.4E6	421.855	402.349
11) T	2,4,5-TP ...	8.110	8.716	6644.9E6	6300.6E6	479.897	633.284 #
12) T	2,4,5-T	8.370	9.093	6505.0E6	4236.6E6	462.061	456.509
13) T	2,4-DB	8.880	9.601	841.7E6	502.7E6	368.281m	429.392
14) T	DINOSEB	9.936	9.929	3410.4E6	2004.9E6	297.344	285.713
15) T	Picloram	9.772	10.905	8327.3E6	5349.6E6	414.100	414.415
16) T	DCPA	10.203	10.860	8980.3E6	5527.5E6	475.099	461.650

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111822\
Data File : PS021198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Nov 2022 13:08
Operator : AR\AJ
Sample : N5598-02MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
FRAC-TANK-0760533MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 00:32:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111622.M
Quant Title : 8080.M
QLast Update : Thu Nov 17 00:46:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

