

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111925\
 Data File : PS032497.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Nov 2025 20:32
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 11/20/2025
 Supervised By :Sohil Jodhani 11/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 20 03:36:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 04 04:59:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.275	7.613	3117.5E6	2881.8E6	687.993m	790.505
Target Compounds							
1) T	Dalapon	2.651	2.601	4645.1E6	4584.6E6	647.054	659.280
2) T	3,5-DICHL...	6.443	6.581	4557.9E6	4134.1E6	704.880	687.995m
3) T	4-Nitroph...	7.095	7.178	959.2E6	1704.7E6	714.501	751.016
5) T	DICAMBA	7.457	7.805	13605.3E6	12806.4E6	721.147	741.437
6) T	MCPD	7.635	7.903	865.3E6	1025.6E6	79.454m	77.106
7) T	MCPA	7.785	8.147	1014.0E6	1249.5E6	75.524	76.635m
8) T	DICHLORPROP	8.163	8.518	3093.6E6	3472.9E6	695.396m	728.209
9) T	2,4-D	8.403	8.859	3222.9E6	3845.7E6	714.334	733.127
10) T	Pentachlo...	8.695	9.355	46851.6E6	43877.1E6	746.769	747.446
11) T	2,4,5-TP ...	9.269	9.741	19140.2E6	19575.2E6	735.835	728.951
12) T	2,4,5-T	9.567	10.168	15471.5E6	17820.8E6	649.021	751.361
13) T	2,4-DB	10.143	10.732	2064.9E6	2759.1E6	783.615	731.594
14) T	DINOSEB	11.334	11.102	14623.3E6	12464.7E6	702.787	704.723
15) T	Picloram	11.171	12.220	15301.3E6	22422.5E6	720.889	603.045
16) T	DCPA	11.634	12.136	24639.9E6	23139.6E6	734.378	726.277

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111925\
Data File : PS032497.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Nov 2025 20:32
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 11/20/2025
Supervised By :Sohil Jodhani 11/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 20 03:36:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110325.M
Quant Title : 8080.M
QLast Update : Tue Nov 04 04:59:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

