

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112023\
 Data File : PS024881.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 17:27
 Operator : AR\AJ
 Sample : 05412-06
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:15:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110323.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 03 15:07:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.365	6.773	824.8E6	220.9E6	404.018	327.509
Target Compounds							
1) T	Dalapon	2.460	2.508	73493333	21106035	30.509	20.033 #
2) T	3,5-DICHL...	0.000	5.981	0	2187770	N.D.	2.043
3) T	4-Nitroph...	6.180	6.413	53665549	6468734	41.402	14.732 #
6) T	MCP	6.683	7.020	47184726	8202744	7.574	4.227 #
7) T	MCPA	6.724f	7.234	14428760	9069172	1.645	2.942 #
8) T	DICHLORPROP	7.126f	7.557	5717242	29965955	2.444	41.021 #
9) T	2,4-D	7.248	7.854	14414443	1270845	5.244	1.420 #
10) T	Pentachlo...	7.520	8.246	21187525	2527298	<MDL	<MDL #
11) T	2,4,5-TP ...	8.102f	8.612	42395203	331648	3.049	<MDL #
12) T	2,4,5-T	8.325	8.968	41542664	5615567	3.177	1.407 #
13) T	2,4-DB	8.818	9.484	79540467	9737886	37.825	19.973 #
14) T	DINOSEB	9.892f	9.823	11956086	9183389	1.212	3.221 #
15) T	Picloram	9.685	10.765	15310699	10846089	<MDL	1.785 #
16) T	DCPA	0.000	10.765	0	10846089	N.D.	2.158



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112023\
Data File : PS024881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 17:27
Operator : AR\AJ
Sample : 05412-06
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:15:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110323.M
Quant Title : 8080.M
QLast Update : Fri Nov 03 15:07:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.460 min
Delta R.T.: -0.021 min
Response: 73493333
Conc: 30.51 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.508 min
Delta R.T.: -0.033 min
Response: 21106035
Conc: 20.03 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 5.732 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 5.981 min
Delta R.T.: 0.017 min
Response: 2187770
Conc: 2.04 ng/ml



#3 4-Nitrophenol


R.T.: 6.180 min
Delta R.T.: -0.016 min
Response: 53665549
Conc: 41.40 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.413 min
Delta R.T.: 0.021 min
Response: 6468734
Conc: 14.73 ng/ml




#4 2,4-DCAA

R.T.: 6.365 min
Delta R.T.: 0.008 min
Response: 824804469
Conc: 404.02 ng/ml



#4 2,4-DCAA


R.T.: 6.773 min
Delta R.T.: -0.004 min
Response: 220937152
Conc: 327.51 ng/ml



#5 DICAMBA


R.T.: 6.529 min
Delta R.T.: 0.031 min
Response: -16168933
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 6.923 min
Delta R.T.: -0.015 min
Response: 3661636
Conc: 1.26 ng/ml




#6 MCPP

R.T.: 6.683 min
Delta R.T.: 0.036 min
Response: 47184726
Conc: 7.57 ug/ml



#6 MCPP


R.T.: 7.020 min
Delta R.T.: -0.011 min
Response: 8202744
Conc: 4.23 ug/ml



#7 MCPA


R.T.: 6.724 min
Delta R.T.: -0.044 min
Response: 14428760
Conc: 1.64 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.234 min
Delta R.T.: -0.003 min
Response: 9069172
Conc: 2.94 ug/ml




#8 DICHLORPROP

R.T.: 7.126 min
Delta R.T.: 0.040 min
Response: 5717242
Conc: 2.44 ng/ml



#8 DICHLORPROP


R.T.: 7.557 min
Delta R.T.: 0.005 min
Response: 29965955
Conc: 41.02 ng/ml



#9 2,4-D


R.T.: 7.248 min
Delta R.T.: -0.036 min
Response: 14414443
Conc: 5.24 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 7.854 min
Delta R.T.: 0.019 min
Response: 1270845
Conc: 1.42 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 8.102 min
Delta R.T.: 0.069 min
Response: 42395203
Conc: 3.05 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.612 min
Delta R.T.: -0.004 min
Response: 331648
Conc: N.D.



#12 2,4,5-T


R.T.: 8.325 min
Delta R.T.: 0.036 min
Response: 41542664
Conc: 3.18 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 8.968 min
Delta R.T.: -0.018 min
Response: 5615567
Conc: 1.41 ng/ml




#13 2,4-DB

R.T.: 8.818 min
Delta R.T.: 0.024 min
Response: 79540467
Conc: 37.83 ng/ml



#13 2,4-DB


R.T.: 9.484 min
Delta R.T.: -0.006 min
Response: 9737886
Conc: 19.97 ng/ml



#14 DINOSEB


R.T.: 9.892 min
Delta R.T.: 0.046 min
Response: 11956086
Conc: 1.21 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 9.823 min
Delta R.T.: 0.007 min
Response: 9183389
Conc: 3.22 ng/ml




#15 Picloram

R.T.: 9.685 min
Delta R.T.: 0.012 min
Response: 15310699
Conc: N.D.



#15 Picloram


R.T.: 10.765 min
Delta R.T.: -0.003 min
Response: 10846089
Conc: 1.79 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 10.111 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.765 min
Delta R.T.: 0.027 min
Response: 10846089
Conc: 2.16 ng/ml