

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112023\
 Data File : PS024882.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 17:51
 Operator : AR\AJ
 Sample : 05412-01MSD (Sig #1); 05412-01MS (Sig #2)
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-1MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/22/2023
 Supervised By :Ankita Jodhani 11/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:16:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110323.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 03 15:07:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.355	6.772	1436.5E6	370.0E6	703.629	548.437m
Target Compounds							
1) T	Dalapon	2.479	2.540	829.2E6	374.9E6	344.220	355.816m
2) T	3,5-DICHL...	5.731	5.962	1316.8E6	429.0E6	419.808	400.662
3) T	4-Nitroph...	6.193	6.388	678.8E6	225.7E6	523.639	514.077
5) T	DICAMBA	6.496	6.933	4220.6E6	1347.6E6	471.703m	463.819
6) T	MCPD	6.646	7.028	200.8E6	70519840	32.233	36.344
7) T	MCPA	6.768	7.234	375.4E6	141.7E6	42.794m	45.976
8) T	DICHLORPROP	7.083	7.548	814.4E6	282.3E6	348.101m	386.499
9) T	2,4-D	7.280	7.830	1304.6E6	427.3E6	474.602	477.417
10) T	Pentachlo...	7.509	8.253	7705.8E6	2307.2E6	236.541	217.111
11) T	2,4,5-TP ...	8.030	8.612	3674.9E6	1080.3E6	264.306	262.844
12) T	2,4,5-T	8.286	8.981	4864.8E6	1424.4E6	372.062	356.977
13) T	2,4-DB	8.792	9.484	619.6E6	132.0E6	294.655	270.671
14) T	DINOSEB	9.843	9.813	403.9E6	104.3E6	40.958	36.565
15) T	Picloram	9.670	10.761	7732.4E6	2604.4E6	434.728	428.634
16) T	DCPA	10.109	10.733	9988.1E6	2788.5E6	583.655	554.843

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112023\
Data File : PS024882.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 17:51
Operator : AR\AJ
Sample : 05412-01MSD (Sig #1); 05412-01MS (Sig #2)
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-1MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:16:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110323.M
Quant Title : 8080.M
QLast Update : Fri Nov 03 15:07:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

