

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112024\
 Data File : PS028573.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2024 18:52
 Operator : AR\AJ
 Sample : P4893-05MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-762MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/21/2024
 Supervised By :Ankita Jodhani 11/21/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 00:02:28 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 13 13:41:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.232	7.733	1303.4E6	574.1E6	509.980m	444.230
Target Compounds							
1) T	Dalapon	2.635	2.699	669.6E6	717.6E6	199.327	296.526m#
2) T	3,5-DICHL...	6.404	6.689	1403.0E6	627.4E6	376.021	324.658
3) T	4-Nitroph...	7.032	7.260	472.6E6	238.7E6	278.241m	276.455
5) T	DICAMBA	7.420	7.933	3937.9E6	1917.1E6	359.885m	336.098
6) T	MCPD	7.600	8.032	230.7E6	138.7E6	32.798m	32.238
7) T	MCPA	7.749	8.277	337.0E6	194.6E6	33.020	31.425
8) T	DICHLORPROP	8.127	8.648	1015.1E6	514.8E6	343.945m	352.432
9) T	2,4-D	8.358	8.979	1295.9E6	652.2E6	405.345m	416.626
10) T	Pentachlo...	8.657	9.507	13394.3E6	5865.3E6	320.928	292.600
11) T	2,4,5-TP ...	9.236	9.884	6481.7E6	3047.5E6	373.142	361.151
12) T	2,4,5-T	9.528	10.304	6398.5E6	2809.5E6	359.981	335.172
13) T	2,4-DB	10.104	10.871	892.4E6	331.0E6	327.417	315.934
14) T	DINOSEB	11.314	11.253	486.7E6	193.2E6	34.138	34.950
15) T	Picloram	11.122	12.344	8127.5E6	2759.7E6	288.620	243.810
16) T	DCPA	11.606	12.294	8485.2E6	3364.5E6	322.467m	346.911m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112024\
Data File : PS028573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2024 18:52
Operator : AR\AJ
Sample : P4893-05MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-762MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/21/2024
Supervised By :Ankita Jodhani 11/21/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 00:02:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
Quant Title : 8080.M
QLast Update : Wed Nov 13 13:41:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

