

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112118\
 Data File : PS002465.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2018 16:55
 Operator : EI\SJ
 Sample : PB115003BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB115003BS

Manual Integrations
 APPROVED

mohammad
 11/23/2018 11:41:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 04:20:27 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 15:54:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.401	10.894	170.1E6	191.0E6	558.585	521.962
Target Compounds							
1) T	Dalapon	2.926	2.372	237.1E6	223.0E6	537.694	518.084
2) T	3,5-DICHL...	9.830	9.050	251.5E6	238.5E6	542.938m	526.113m
3) T	4-Nitroph...	10.994	10.041	94807126	122.5E6	584.315	699.035
5) T	DICAMBA	11.707	11.177	705.6E6	713.4E6	509.231m	527.321m
6) T	MCPD	12.047	11.405	61013821	53419149	65.010m	58.889m
7) T	MCPA	12.295	11.780	78955157	83407547	54.665m	59.271m
8) T	DICHLORPROP	12.895	12.300	180.5E6	187.5E6	521.395	565.641m
9) T	2,4-D	13.254	12.770	215.9E6	243.8E6	671.841m	610.500m
10) T	Pentachlo...	13.631	13.316	2706.3E6	2885.7E6	559.486	567.042m
11) T	2,4,5-TP ...	14.500	13.934	1032.6E6	1135.2E6	549.506	562.739
12) T	2,4,5-T	14.897	14.498	847.8E6	889.5E6	502.978	568.997
13) T	2,4-DB	15.644	15.148	112.2E6	133.0E6	579.132m	586.549m
14) T	DINOSEB	17.118	15.575	712.0E6	752.7E6	498.562	502.943
15) T	Picloram	16.890	16.752	1008.7E6	1053.9E6	474.959	396.162m
16) T	DCPA	17.462	16.806	780.7E6	1231.1E6	337.013m	513.352m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112118\
Data File : PS002465.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2018 16:55
Operator : EI\SJ
Sample : PB115003BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB115003BS

Manual Integrations
APPROVED

mohammad
11/23/2018 11:41:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 04:20:27 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 15:54:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

