

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112123\
 Data File : PS024892.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 11:35
 Operator : AR\AJ
 Sample : 05434-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 22:53:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110323.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 03 15:07:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.356	6.773	686.1E6	224.4E6	336.066	332.679
Target Compounds							
1) T	Dalapon	2.445	2.504	24091963	6506481	10.001	6.176 #
3) T	4-Nitroph...	6.189	6.418	39647479	12649912	30.587	28.809
5) T	DICAMBA	0.000	6.922	0	628506	N.D.	<MDL
6) T	MCP	6.674	7.050	39917054	3077819	6.407	1.586 #
7) T	MCPA	6.802	7.234	16114717	12253856	1.837	3.975 #
8) T	DICHLORPROP	7.038f	7.557	8085195	18881988	3.456	25.848 #
9) T	2,4-D	7.245	7.846	17906374	2602181	6.514	2.907 #
10) T	Pentachlo...	7.508	8.246	19717227	1420685	<MDL	<MDL #
12) T	2,4,5-T	8.302	8.983	138.6E6	1701598	10.597	<MDL #
13) T	2,4-DB	8.807	9.490	132.1E6	19463274	62.839	39.920 #
14) T	DINOSEB	9.788f	9.801	389.7E6	32043262	39.520	11.239 #
15) T	Picloram	9.674	10.773	103.2E6	55669234	5.803	9.162 #
16) T	DCPA	0.000	10.773	0	55669234	N.D.	11.077



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112123\
Data File : PS024892.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 11:35
Operator : AR\AJ
Sample : 05434-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 22:53:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110323.M
Quant Title : 8080.M
QLast Update : Fri Nov 03 15:07:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.445 min
Delta R.T.: -0.036 min
Response: 24091963
Conc: 10.00 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.504 min
Delta R.T.: -0.037 min
Response: 6506481
Conc: 6.18 ng/ml




#3 4-Nitrophenol

R.T.: 6.189 min
Delta R.T.: -0.008 min
Response: 39647479
Conc: 30.59 ng/ml



#3 4-Nitrophenol


R.T.: 6.418 min
Delta R.T.: 0.026 min
Response: 12649912
Conc: 28.81 ng/ml



#4 2,4-DCAA


R.T.: 6.356 min
Delta R.T.: -0.001 min
Response: 686079314
Conc: 336.07 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.773 min
Delta R.T.: -0.004 min
Response: 224424569
Conc: 332.68 ng/ml




#6 MCPP

R.T.: 6.674 min
Delta R.T.: 0.027 min
Response: 39917054
Conc: 6.41 ug/ml



#6 MCPP


R.T.: 7.050 min
Delta R.T.: 0.019 min
Response: 3077819
Conc: 1.59 ug/ml



#7 MCPA


R.T.: 6.802 min
Delta R.T.: 0.034 min
Response: 16114717
Conc: 1.84 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.234 min
Delta R.T.: -0.003 min
Response: 12253856
Conc: 3.97 ug/ml




#8 DICHLORPROP

R.T.: 7.038 min
Delta R.T.: -0.047 min
Response: 8085195
Conc: 3.46 ng/ml



#8 DICHLORPROP


R.T.: 7.557 min
Delta R.T.: 0.006 min
Response: 18881988
Conc: 25.85 ng/ml



#9 2,4-D


R.T.: 7.245 min
Delta R.T.: -0.039 min
Response: 17906374
Conc: 6.51 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 7.846 min
Delta R.T.: 0.011 min
Response: 2602181
Conc: 2.91 ng/ml




#12 2,4,5-T

R.T.: 8.302 min
Delta R.T.: 0.013 min
Response: 138552762
Conc: 10.60 ng/ml



#12 2,4,5-T


R.T.: 8.983 min
Delta R.T.: -0.003 min
Response: 1701598
Conc: N.D.



#13 2,4-DB


R.T.: 8.807 min
Delta R.T.: 0.013 min
Response: 132141213
Conc: 62.84 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 9.490 min
Delta R.T.: 0.000 min
Response: 19463274
Conc: 39.92 ng/ml




#14 DINOSEB

R.T.: 9.788 min
Delta R.T.: -0.057 min
Response: 389748774
Conc: 39.52 ng/ml



#14 DINOSEB


R.T.: 9.801 min
Delta R.T.: -0.015 min
Response: 32043262
Conc: 11.24 ng/ml



#15 Picloram


R.T.: 9.674 min
Delta R.T.: 0.002 min
Response: 103220477
Conc: 5.80 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 10.773 min
Delta R.T.: 0.006 min
Response: 55669234
Conc: 9.16 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 10.111 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 10.773 min
Delta R.T.: 0.035 min
Response: 55669234
Conc: 11.08 ng/ml